

Certificate of Analysis



Level 2, 321 Kent Street, Sydney NSW 2000, Australia

Phone: +61 2 8228 2719

Email: info@geneabiomedx.com

Website: geneabiomedx.com

Product Details

Product description:	Oocyte Retrieval Buffer
Product code	ORB-50
Lot number:	J000008032
Expiry date:	28 Jun. 2027

Intended Use

Gems Oocyte Retrieval Buffer is used in the removal of oocytes from ovarian follicles.

Caution: Federal law (USA) restricts this device to sale by or on the order of a physician or practitioner trained in its use.

Product Analysis

Assay	Specification	Result
pH at 25 °C in air	7.3 – 7.5	Pass
Osmolality	280 – 290 mOsm/kg	Pass
MEA	MEA (1-cell): ≥80% developed to blastocyst stage or beyond by 96 hours	Pass
Endotoxin	<0.40 EU/mL	Pass
Sterility	No Growth	Pass
Gentamicin	0.0075 – 0.0125 mg/mL	Pass

Product passed and released by Quality

Signature:

A handwritten signature in blue ink, appearing to read 'C. De Cruz'.

Electronically signed by: Caitlin De Cruz
Reason: QA Approval
Date: Aug 5, 2026 19:02:59 GMT
+10

Name: Caitlin De Cruz

Job Title: Quality Assurance Associate

Date: 05 Aug. 2026

Certificate of Analysis - J000008032

Final Audit Report

2026-08-05

Created:	2026-08-05 (Australian Eastern Standard Time)
By:	Caitlin De Cruz (caitlin.decruz@geneabiomedx.com)
Status:	Signed
Transaction ID:	CBJCHBCAABAAysTOFIHC7zh81yVrnSUtv9Cw60yfVfWp

"Certificate of Analysis - J000008032" History

-  Document created by Caitlin De Cruz (caitlin.decruz@geneabiomedx.com)
2026-08-05 - 7:01:47 PM GMT+10- IP address: 157.211.44.217
-  Document emailed to Caitlin De Cruz (caitlin.decruz@geneabiomedx.com) for signature
2026-08-05 - 7:02:01 PM GMT+10
-  Email viewed by Caitlin De Cruz (caitlin.decruz@geneabiomedx.com)
2026-08-05 - 7:02:23 PM GMT+10- IP address: 48.215.3.159
-  Caitlin De Cruz (caitlin.decruz@geneabiomedx.com) authenticated with Adobe Acrobat Sign.
Challenge: The user opened the agreement.
2026-08-05 - 7:02:25 PM GMT+10
-  Caitlin De Cruz (caitlin.decruz@geneabiomedx.com) authenticated with Adobe Acrobat Sign.
Challenge: The user completed the signing ceremony.
2026-08-05 - 7:02:57 PM GMT+10
-  Document e-signed by Caitlin De Cruz (caitlin.decruz@geneabiomedx.com)
Signing reason: QA Approval
Signature Date: 2026-08-05 - 7:02:59 PM GMT+10 - Time Source: server- IP address: 157.211.44.217 - Signature Appearance Selected: IMAGE
-  Agreement completed.
2026-08-05 - 7:02:59 PM GMT+10