

Certificate of Analysis



Level 2, 321 Kent Street, Sydney NSW 2000, Australia

Phone: +61 2 8228 2719
Email: info@geneabiomedx.com
Website: geneabiomedx.com

Product Details

Product description: VitBase
Product code VBS-20
Lot number: J000008040
Expiry date: 29 Jun. 2027

Intended Use

VitBase is a HEPES buffered medium that is used to maintain embryos for a short period of time in a non-gassed environment.

USA only: Gems VitBase is a HEPES buffered medium that is used to maintain embryos for a short period of time in a non-gassed environment during embryo vitrification and warming procedures. The embryos can be placed in this medium for a maximum of 10 minutes.

Caution: Federal law (USA) restricts this device to sale by or on the order of a physician or practitioner trained in its use.

Product Analysis

Assay	Specification	Result
pH at 25 °C in air	7.3 – 7.5	Pass
Osmolality	295 – 305 mOsm/kg	Pass
MEA	MEA (1-cell): ≥80% developed to blastocyst stage or beyond by 96 hours	Pass
Endotoxin	<0.40 EU/mL	Pass
Sterility	No Growth	Pass
HSA Analysis	15.00 – 25.00 mg/mL	Pass

Product passed and released by Quality

Signature:

Electronically signed by: Nicole Connor
Reason: QA Approval
Date: Aug 5, 2026 11:57:06 GMT +10

Name: Nicole Connor

Job Title: Quality Assurance Associate

Date: 05 Aug. 2026

Certificate of Analysis- J000008040

Final Audit Report

2026-08-05

Created:	2026-08-05 (Australian Eastern Standard Time)
By:	Nicole Connor (nicole.connor@geneabiomedx.com)
Status:	Signed
Transaction ID:	CBJCHBCAABAA1HXKfqeUKFPL60ZKaslPwjJAgUjd2Hu

"Certificate of Analysis- J000008040" History

-  Document created by Nicole Connor (nicole.connor@geneabiomedx.com)
2026-08-05 - 11:55:58 AM GMT+10- IP address: 149.167.36.126
-  Document emailed to Nicole Connor (nicole.connor@geneabiomedx.com) for signature
2026-08-05 - 11:56:15 AM GMT+10
-  Email viewed by Nicole Connor (nicole.connor@geneabiomedx.com)
2026-08-05 - 11:56:22 AM GMT+10- IP address: 48.215.3.159
-  Nicole Connor (nicole.connor@geneabiomedx.com) authenticated with Adobe Acrobat Sign.
Challenge: The user opened the agreement.
2026-08-05 - 11:56:34 AM GMT+10
-  Nicole Connor (nicole.connor@geneabiomedx.com) authenticated with Adobe Acrobat Sign.
Challenge: The user completed the signing ceremony.
2026-08-05 - 11:57:04 AM GMT+10
-  Document e-signed by Nicole Connor (nicole.connor@geneabiomedx.com)
Signing reason: QA Approval
Signature Date: 2026-08-05 - 11:57:06 AM GMT+10 - Time Source: server- IP address: 149.167.36.126 - Signature Appearance Selected: DRAW
-  Agreement completed.
2026-08-05 - 11:57:06 AM GMT+10