

Certificate of Analysis



Level 2, 321 Kent Street, Sydney NSW 2000, Australia

Phone: +61 2 8228 2719
Email: info@geneabiomedx.com
Website: geneabiomedx.com

Product Details

Product description: Sperm Buffer
Product code SPB-50
Lot number: J000006797
Expiry date: 28 Jul. 2025

Intended Use

Sperm Buffer is used to wash and resuspend sperm for the insemination step in intrauterine insemination (IUI), IVF or in diagnostic washing.

Product Analysis

Assay	Specification	Result
pH at 25 °C in air	7.3 – 7.5	Pass
Osmolality	285 – 295 mOsm/kg	Pass
MEA	MEA (1-cell): ≥80% developed to blastocyst stage or beyond by 96 hours	Pass
Sperm Toxicity Test	Sperm Survival Tested	Pass
Endotoxin	<0.40 EU/mL	Pass
Sterility	No Growth	Pass
HSA Analysis	7.50 – 12.50 mg/mL	Pass
Gentamicin	0.0075 – 0.0125 mg/mL	Pass

Product passed and released by Quality

Signature:

Electronically signed by: Lalitha Kalyanaraman
Reason: QA Approval
Date: Sep 4, 2024 13:12 GMT+10

Name: Lalitha Kalyanaraman

Job Title: QA Associate

Date: 04 Sep. 2024

Certificate of Analysis - J000006797

Final Audit Report

2024-09-04

Created:	2024-09-04 (Australian Eastern Standard Time)
By:	Lalitha Kalyanaraman (lalitha.kalyanaraman@geneabiomedx.com)
Status:	Signed
Transaction ID:	CBJCHBCAABAAfMAWLpNRpudR9u3wZJkbAa-0QLIWM8E7

"Certificate of Analysis - J000006797" History

-  Document created by Lalitha Kalyanaraman (lalitha.kalyanaraman@geneabiomedx.com)
2024-09-04 - 1:09:36 PM GMT+10- IP address: 58.111.90.210
-  Document emailed to Lalitha Kalyanaraman (lalitha.kalyanaraman@geneabiomedx.com) for signature
2024-09-04 - 1:09:55 PM GMT+10
-  Lalitha Kalyanaraman (lalitha.kalyanaraman@geneabiomedx.com) authenticated with Adobe Acrobat Sign.
Challenge: The user opened the agreement.
2024-09-04 - 1:10:33 PM GMT+10
-  Document e-signed by Lalitha Kalyanaraman (lalitha.kalyanaraman@geneabiomedx.com)
Signing reason: QA Approval
Signature Date: 2024-09-04 - 1:12:30 PM GMT+10 - Time Source: server- IP address: 58.111.90.210
-  Agreement completed.
2024-09-04 - 1:12:30 PM GMT+10
-  Lalitha Kalyanaraman (lalitha.kalyanaraman@geneabiomedx.com) authenticated with Adobe Acrobat Sign.
Challenge: The user completed the signing ceremony.
2024-09-04 - 1:12:30 PM GMT+10