

Certificate of Analysis



Level 2, 321 Kent Street, Sydney NSW 2000, Australia

Phone:

+61 2 8228 2719

Email:

info@geneabiomedx.com

Website:

geneabiomedx.com

Product Details

Product description: Oocyte Retrieval Buffer

Product code ORB-50

Lot number: J000006943

Expiry date: 29 Sep. 2025

Intended Use

Gems Oocyte Retrieval Buffer is used in the removal of oocytes from ovarian follicles.

Caution: Federal law (USA) restricts this device to sale by or on the order of a physician or practitioner trained in its use.

Product Analysis

| Assay | Specification | Result |
|--------------------|--|--------|
| pH at 25 °C in air | 7.3 – 7.5 | Pass |
| Osmolality | 280 – 290 mOsm/kg | Pass |
| MEA | MEA (1-cell): ≥80% developed to blastocyst stage or beyond by 96 hours | Pass |
| Endotoxin | <0.40 EU/mL | Pass |
| Sterility | No Growth | Pass |
| Gentamicin | 0.0075 – 0.0125 mg/mL | Pass |

Product passed and released by Quality

Signature:

A handwritten signature in black ink, appearing to read 'E. McGregor'.

Name:

Eila McGregor

Job Title:

Quality Assurance Associate

Date:

18 Nov. 2024