

# Certificate of Analysis



Level 2, 321 Kent Street, Sydney NSW 2000, Australia

Phone:

+61 2 8228 2719

Email:

info@geneabiomedx.com

Website:

geneabiomedx.com

## Product Details

**Product description:** Sperm Wash Gradient Set  
**Product code** SWG-01  
**Lot number:** J000006978  
**Expiry date:** 30 Sep. 2025

## Intended Use

Sperm Wash Gradient Set is used to separate sperm from seminal plasma as well as separating highly motile sperm in preparation for insemination.

## Product Analysis

Assay	Specification		Result
	45% gradient	90% gradient	
pH at 25 °C in air	7.3 – 7.5		Pass
Osmolality	285 - 295 mOsm/kg	280 - 290 mOsm/kg	Pass
Sperm Toxicity Test	Sperm Survival Tested		Pass
Endotoxin	<0.40 EU/mL		Pass
Sterility	No Growth		Pass
HSA Analysis	7.50 – 12.50 mg/mL		Pass
Gentamicin	0.0075 – 0.0125 mg/mL		Pass

**Product passed and released by Quality**

**Signature:**

A handwritten signature in black ink, appearing to read 'Eila McGregor'.

**Name:**

Eila McGregor

**Job Title:**

Quality Assurance Associate

**Date:**

14 Nov. 2024