

Certificate of Analysis



Level 2, 321 Kent Street, Sydney NSW 2000, Australia

Phone:

+61 2 8228 2719

Email:

info@geneabiomedx.com

Website:

geneabiomedx.com

Product Details

Product description: Oocyte Retrieval Buffer

Product code ORB-50

Lot number: J000007002

Expiry date: 10 Oct. 2025

Intended Use

Gems Oocyte Retrieval Buffer is used in the removal of oocytes from ovarian follicles.

Caution: Federal law (USA) restricts this device to sale by or on the order of a physician or practitioner trained in its use.

Product Analysis

Assay	Specification	Result
pH at 25 °C in air	7.3 – 7.5	Pass
Osmolality	280 – 290 mOsm/kg	Pass
MEA	MEA (1-cell): ≥80% developed to blastocyst stage or beyond by 96 hours	Pass
Endotoxin	<0.40 EU/mL	Pass
Sterility	No Growth	Pass
Gentamicin	0.0075 – 0.0125 mg/mL	Pass

Product passed and released by Quality

Signature:

A handwritten signature in black ink that reads 'E. McGregor'.

Name:

Eila McGregor

Job Title:

Quality Assurance Associate

Date:

04 Dec. 2024