

Certificate of Analysis



Level 2, 321 Kent Street, Sydney NSW 2000, Australia

Phone: +61 2 8228 2719
Email: info@geneabiomedx.com
Website: geneabiomedx.com

Product Details

Product description: Sperm Wash Gradient Set
Product code SWG-01
Lot number: J000007069
Expiry date: 10 Nov. 2025

Intended Use

Sperm Wash Gradient Set is used to separate sperm from seminal plasma as well as separating highly motile sperm in preparation for insemination.

Product Analysis

Assay	Specification		Result
	45% gradient	90% gradient	
pH at 25 °C in air	7.3 – 7.5		Pass
Osmolality	285 - 295 mOsm/kg	280 - 290 mOsm/kg	Pass
Sperm Toxicity Test	Sperm Survival Tested		Pass
Endotoxin	<0.40 EU/mL		Pass
Sterility	No Growth		Pass
HSA Analysis	7.50 – 12.50 mg/mL		Pass
Gentamicin	0.0075 – 0.0125 mg/mL		Pass

Product passed and released by Quality

Signature:

A handwritten signature in black ink, appearing to read 'Elliot Atchison'.

*Electronically signed by:
Elliot Atchison
Reason: Approval
Date: Apr 9, 2025 13:55
GMT+10*

Name: Elliot Atchison

Job Title: QA Lead

Date: 09 Apr. 2025

QFRM221 GEMS Certificate of Analysis








SWG-01 Sperm Wash Gradient Set

Final Audit Report

2025-04-09

Created:	2025-04-09 (Australian Eastern Standard Time)
By:	Elliot Atchison (elliott.atchison@geneabiomedx.com)
Status:	Approved
Transaction ID:	CBJCHBCAABAARZkMo2Mw50_tuH36C8sA18WxYy28tzx0

"QFRM221 GEMS Certificate of Analysis SWG-01 Sperm Wash Gradient Set" History

-  Document created by Elliot Atchison (elliott.atchison@geneabiomedx.com)
2025-04-09 - 1:54:02 PM GMT+10- IP address: 202.67.82.173
-  Document emailed to Elliot Atchison (elliott.atchison@geneabiomedx.com) for approval
2025-04-09 - 1:54:25 PM GMT+10
-  Email viewed by Elliot Atchison (elliott.atchison@geneabiomedx.com)
2025-04-09 - 1:54:36 PM GMT+10- IP address: 104.210.79.75
-  Elliot Atchison (elliott.atchison@geneabiomedx.com) authenticated with Adobe Acrobat Sign.
Challenge: The user opened the agreement.
2025-04-09 - 1:54:42 PM GMT+10
-  Elliot Atchison (elliott.atchison@geneabiomedx.com) authenticated with Adobe Acrobat Sign.
Challenge: The user completed the signing ceremony.
2025-04-09 - 1:55:07 PM GMT+10
-  Document approved by Elliot Atchison (elliott.atchison@geneabiomedx.com)
Signing reason: Approval
Approval Date: 2025-04-09 - 1:55:08 PM GMT+10 - Time Source: server- IP address: 202.67.82.173
-  Agreement completed.
2025-04-09 - 1:55:08 PM GMT+10