Certificate of Analysis



Phone: Email: Website:

+61 2 8228 2719 info@geneabiomedx.com geneabiomedx.com

Product Details

Product description:	Blastocyst Medium
Product code	BLM-50
Lot number:	J000007234
Expiry date:	23 Feb. 2026

Intended Use

Blastocyst Medium is used to provide necessary nutrients for embryo development *in vitro* from the cleavage stage to the blastocyst stage. The embryos may remain in this solution for up to 3 days prior to being transferred to the uterus.

USA only: Gems Blastocyst Medium is for *in vitro* culture of embryos from the cleavage stage to the blastocyst stage of development. This medium can be used for transfer of blastocyst stage embryos into the uterus.

Caution: Federal law (USA) restricts this device to sale by or on the order of a physician or practitioner trained in its use.

Product Analysis

Assay	Specification	Result
pH at 37°C with 6 % CO₂	7.25 – 7.45	Pass
Osmolality	285 – 295 mOsm/kg	Pass
MEA	MEA (1-cell): ≥80% developed to blastocyst stage or beyond by 96 hours	Pass
Endotoxin	<0.40 EU/mL	Pass
Sterility	No Growth	Pass
HSA Analysis	3.75 - 6.25 mg/mL	Pass
Gentamicin	0.0075 – 0.0125 mg/mL	Pass

Product passed and released by Quality

Signature:	Emgregation Control Co
Name:	Eila McGregor
Job Title:	Quality Assurance Associate
Date:	29 Apr. 2025

Certificate of Analysis - J000007234

Final Audit Report

2025-04-29

Created:	2025-04-29 (Australian Eastern Standard Time)
Ву:	Eila McGregor (eila.mcgregor@geneabiomedx.com)
Status:	Signed
Transaction ID:	CBJCHBCAABAAVsYTYsQQ-3Tbpl68oX3F_rrFz3THYequ

"Certificate of Analysis - J000007234" History

- Document created by Eila McGregor (eila.mcgregor@geneabiomedx.com) 2025-04-29 12:26:34 PM GMT+10- IP address: 220.245.114.106
- Document emailed to Eila McGregor (eila.mcgregor@geneabiomedx.com) for signature 2025-04-29 - 12:26:49 PM GMT+10
- Email viewed by Eila McGregor (eila.mcgregor@geneabiomedx.com) 2025-04-29 - 12:26:56 PM GMT+10- IP address: 4.199.143.157
- Eila McGregor (eila.mcgregor@geneabiomedx.com) authenticated with Adobe Acrobat Sign. Challenge: The user opened the agreement.
 2025-04-29 - 12:27:05 PM GMT+10
- Document e-signed by Eila McGregor (eila.mcgregor@geneabiomedx.com) Signing reason: QA Approval Signature Date: 2025-04-29 - 12:27:28 PM GMT+10 - Time Source: server- IP address: 220.245.114.106
- Agreement completed.
 2025-04-29 12:27:28 PM GMT+10
- Eila McGregor (eila.mcgregor@geneabiomedx.com) authenticated with Adobe Acrobat Sign. Challenge: The user completed the signing ceremony.
 2025-04-29 - 12:27:28 PM GMT+10

