

Certificate of Analysis



Level 2, 321 Kent Street, Sydney NSW 2000, Australia

Phone: +61 2 8228 2719

Email: info@geneabiomedx.com

Website: geneabiomedx.com

Product Details

Product description:	GERI Medium
Product code	ONE-20
Lot number:	J000007215
Expiry date:	05 Feb. 2026

Intended Use

Geri Medium is used to provide necessary nutrients for embryo development *in vitro*. The embryos can remain in this solution for up to 5 days prior to being transferred to the uterus.

USA only: Gems Geri Medium is for *in vitro* culture of embryos from fertilisation to the blastocyst stage of development. This medium can be used for transfer of embryos into the uterus.

Caution: Federal law (USA) restricts this device to sale by or on the order of a physician or practitioner trained in its use.

Product Analysis

Assay	Specification	Result
pH at 37 °C with 6 % CO ₂	7.3 – 7.5	Pass
Osmolality	285 – 295 mOsm/kg	Pass
MEA	MEA (1-cell): ≥80% developed to blastocyst stage or beyond by 96 hours	Pass
Endotoxin	<0.40 EU/mL	Pass
Sterility	No Growth	Pass
HSA Analysis	3.75 – 6.25 mg/mL	Pass
Gentamicin	0.0075 – 0.0125 mg/mL	Pass

Product passed and released by Quality

Signature:

Electronically signed by: Eila
McGregor
Reason: QA Approval
Date: May 20, 2025 12:23 GMT+10

Name: Eila McGregor

Job Title: Quality Assurance Associate

Date: 20 May. 2025








Certificate of Analysis - J000007215

Final Audit Report

2025-05-20

Created:	2025-05-20 (Australian Eastern Standard Time)
By:	Eila McGregor (eila.mcgregor@geneabiomedx.com)
Status:	Signed
Transaction ID:	CBJCHBCAABAAwJmZHITPDeBPxB3cHOY8m7frcTBJpCR

"Certificate of Analysis - J000007215" History

-  Document created by Eila McGregor (eila.mcgregor@geneabiomedx.com)
2025-05-20 - 12:22:39 PM GMT+10- IP address: 163.47.70.127
-  Document emailed to Eila McGregor (eila.mcgregor@geneabiomedx.com) for signature
2025-05-20 - 12:23:10 PM GMT+10
-  Email viewed by Eila McGregor (eila.mcgregor@geneabiomedx.com)
2025-05-20 - 12:23:19 PM GMT+10- IP address: 48.215.4.33
-  Eila McGregor (eila.mcgregor@geneabiomedx.com) authenticated with Adobe Acrobat Sign.
Challenge: The user opened the agreement.
2025-05-20 - 12:23:27 PM GMT+10
-  Eila McGregor (eila.mcgregor@geneabiomedx.com) authenticated with Adobe Acrobat Sign.
Challenge: The user completed the signing ceremony.
2025-05-20 - 12:23:45 PM GMT+10
-  Document e-signed by Eila McGregor (eila.mcgregor@geneabiomedx.com)
Signing reason: QA Approval
Signature Date: 2025-05-20 - 12:23:46 PM GMT+10 - Time Source: server- IP address: 163.47.70.127
-  Agreement completed.
2025-05-20 - 12:23:46 PM GMT+10