

Certificate of Analysis



Level 2, 321 Kent Street, Sydney NSW 2000, Australia

Phone: +61 2 9229 6420

Email: info@geneabiomedx.com

Website: geneabiomedx.com

Product Details

Product Description	GERI DISH
Product Code	GERI-DSH-20
Lot Number	GD0044-001
Manufacture Date	2023-03
Expiry date	2026-02

Product Specification

Requirement	Specification	Result
MEA	$\geq 80\%$ of 2PN's developed to Blastocyst stage by 96 hrs	Pass
Endotoxin	< 20EU/Device	Pass
Sterility	No Growth	Pass

This product complies with approved and released procedures within the Genea Biomedx EN/ISO 13485 compliant Quality Management System. This product and all its identified Part Numbers and Revisions conform to approved product specifications retained within the QMS. Where cited, lot numbers indicate traceable items within the QMS.

Product Passed and Released by Quality Assurance

Signature: 

Name: Eila McGregor

Job Title: Quality Assurance Co-ordinator

Date: 31 Aug. 2023