

# Certificate of Analysis



Level 2, 321 Kent Street, Sydney NSW 2000, Australia

Phone: +61 2 8228 2719  
Email: [info@geneabiomedx.com](mailto:info@geneabiomedx.com)  
Website: [geneabiomedx.com](http://geneabiomedx.com)

## Product Details

**Product description:** Sperm Buffer  
**Product code** SPB-50  
**Lot number:** J000007337  
**Expiry date:** 07 May. 2026

## Intended Use

Sperm Buffer is used to wash and resuspend sperm for the insemination step in intrauterine insemination (IUI), IVF or in diagnostic washing.

## Product Analysis

Assay	Specification	Result
pH at 25 °C in air	7.3 – 7.5	Pass
Osmolality	285 – 295 mOsm/kg	Pass
MEA	MEA (1-cell): ≥80% developed to blastocyst stage or beyond by 96 hours	Pass
Sperm Toxicity Test	Sperm Survival Tested	Pass
Endotoxin	<0.40 EU/mL	Pass
Sterility	No Growth	Pass
HSA Analysis	7.50 – 12.50 mg/mL	Pass
Gentamicin	0.0075 – 0.0125 mg/mL	Pass

## Product passed and released by Quality

**Signature:**

A handwritten signature in black ink, appearing to read 'Eila McGregor'.

Electronically signed by:  
Eila McGregor  
Reason: QA Approval  
Date: Jun 23, 2025 08:21  
GMT+10

**Name:** Eila McGregor  
**Job Title:** Quality Assurance Associate  
**Date:** 23 Jun. 2025








# Certificate of Analysis - J000007337

Final Audit Report

2025-06-23

Created:	2025-06-23 (Australian Eastern Standard Time)
By:	Eila McGregor (eila.mcgregor@geneabiomedx.com)
Status:	Signed
Transaction ID:	CBJCHBCAABAApuy0gRfuER3HxHhamQhZoBQEVc9CqDy4

## "Certificate of Analysis - J000007337" History

-  Document created by Eila McGregor (eila.mcgregor@geneabiomedx.com)  
2025-06-23 - 8:20:41 AM GMT+10- IP address: 220.245.114.106
-  Document emailed to Eila McGregor (eila.mcgregor@geneabiomedx.com) for signature  
2025-06-23 - 8:20:59 AM GMT+10
-  Email viewed by Eila McGregor (eila.mcgregor@geneabiomedx.com)  
2025-06-23 - 8:21:05 AM GMT+10- IP address: 48.215.3.140
-  Eila McGregor (eila.mcgregor@geneabiomedx.com) authenticated with Adobe Acrobat Sign.  
Challenge: The user opened the agreement.  
2025-06-23 - 8:21:17 AM GMT+10
-  Document e-signed by Eila McGregor (eila.mcgregor@geneabiomedx.com)  
Signing reason: QA Approval  
Signature Date: 2025-06-23 - 8:21:43 AM GMT+10 - Time Source: server- IP address: 220.245.114.106
-  Agreement completed.  
2025-06-23 - 8:21:43 AM GMT+10
-  Eila McGregor (eila.mcgregor@geneabiomedx.com) authenticated with Adobe Acrobat Sign.  
Challenge: The user completed the signing ceremony.  
2025-06-23 - 8:21:43 AM GMT+10