

Certificate of Analysis



Level 2, 321 Kent Street, Sydney NSW 2000, Australia

Phone: +61 2 8228 2719
Email: info@geneabiomedx.com
Website: geneabiomedx.com

Product Details

Product description: Sperm Medium
Product code SPM-50
Lot number: J000007434
Expiry date: 02 Jul. 2026

Intended Use

Sperm Medium is used to wash and resuspend sperm for the insemination step in IUI, IVF or in diagnostic washing, optimised for storage in a 6% CO₂ incubator.

Product Analysis

Assay	Specification	Result
pH at 37°C with 6 % CO ₂	7.3 – 7.5	Pass
Osmolality	285 – 295 mOsm/kg	Pass
MEA	MEA (1-cell): ≥80% developed to blastocyst stage or beyond by 96 hours	Pass
Sperm Toxicity Test	Sperm Survival Tested	Pass
Endotoxin	<0.40 EU/mL	Pass
Sterility	No Growth	Pass
HSA Analysis	7.50 – 12.50 mg/mL	Pass
Gentamicin	0.0075 – 0.0125 mg/mL	Pass

Product passed and released by Quality

Signature:

Name: Eila McGregor

Job Title: Quality Assurance Associate

Date: 05 Aug. 2025