

Certificate of Analysis



Level 2, 321 Kent Street, Sydney NSW 2000, Australia

Phone: +61 2 8228 2719
Email: info@geneabiomedx.com
Website: geneabiomedx.com

Product Details

Product description: GERI Medium
Product code ONE-50
Lot number: J000007425
Expiry date: 30 Jun. 2026

Intended Use

Geri Medium is used to provide necessary nutrients for embryo development *in vitro*. The embryos can remain in this solution for up to 5 days prior to being transferred to the uterus.

USA only: Gems Geri Medium is for *in vitro* culture of embryos from fertilisation to the blastocyst stage of development. This medium can be used for transfer of embryos into the uterus.

Caution: Federal law (USA) restricts this device to sale by or on the order of a physician or practitioner trained in its use.

Product Analysis

Assay	Specification	Result
pH at 37 °C with 6 % CO ₂	7.3 – 7.5	Pass
Osmolality	285 – 295 mOsm/kg	Pass
MEA	MEA (1-cell): ≥80% developed to blastocyst stage or beyond by 96 hours	Pass
Endotoxin	<0.40 EU/mL	Pass
Sterility	No Growth	Pass
HSA Analysis	3.75 – 6.25 mg/mL	Pass
Gentamicin	0.0075 – 0.0125 mg/mL	Pass

Product passed and released by Quality

Signature:

Electronically signed by: Daire Dowling
Reason: QA Approval
Date: Aug 11, 2025 08:31:25 GMT+10

Name: Daire Dowling

Job Title: Quality Assurance Associate

Date: 11 Aug. 2025








Certificate of Analysis -J000007425

Final Audit Report

2025-08-11

Created:	2025-08-11 (Australian Eastern Standard Time)
By:	Daire Dowling (daire.dowling@geneabiomedx.com)
Status:	Signed
Transaction ID:	CBJCHBCAABAAVM5hZqRjDSnsf2o4g-ZL8SQcQdldDd8t

"Certificate of Analysis -J000007425" History

-  Document created by Daire Dowling (daire.dowling@geneabiomedx.com)
2025-08-11 - 8:30:08 AM GMT+10- IP address: 120.17.113.48
-  Document emailed to Daire Dowling (daire.dowling@geneabiomedx.com) for signature
2025-08-11 - 8:30:33 AM GMT+10
-  Email viewed by Daire Dowling (daire.dowling@geneabiomedx.com)
2025-08-11 - 8:30:39 AM GMT+10- IP address: 68.218.165.164
-  Daire Dowling (daire.dowling@geneabiomedx.com) authenticated with Adobe Acrobat Sign.
Challenge: The user opened the agreement.
2025-08-11 - 8:30:53 AM GMT+10
-  Daire Dowling (daire.dowling@geneabiomedx.com) authenticated with Adobe Acrobat Sign.
Challenge: The user completed the signing ceremony.
2025-08-11 - 8:31:24 AM GMT+10
-  Document e-signed by Daire Dowling (daire.dowling@geneabiomedx.com)
Signing reason: QA Approval
Signature Date: 2025-08-11 - 8:31:25 AM GMT+10 - Time Source: server- IP address: 120.17.113.48
-  Agreement completed.
2025-08-11 - 8:31:25 AM GMT+10