

Certificate of Analysis



Level 2, 321 Kent Street, Sydney NSW 2000, Australia

Phone: +61 2 8228 2719
Email: info@geneabiomedx.com
Website: geneabiomedx.com

Product Details

Product description: Sperm Medium
Product code SPM-50
Lot number: J000007458
Expiry date: 15 Jul. 2026

Intended Use

Sperm Medium is used to wash and resuspend sperm for the insemination step in IUI, IVF or in diagnostic washing, optimised for storage in a 6% CO₂ incubator.

Product Analysis

Assay	Specification	Result
pH at 37°C with 6 % CO ₂	7.3 – 7.5	Pass
Osmolality	285 – 295 mOsm/kg	Pass
MEA	MEA (1-cell): ≥80% developed to blastocyst stage or beyond by 96 hours	Pass
Sperm Toxicity Test	Sperm Survival Tested	Pass
Endotoxin	<0.40 EU/mL	Pass
Sterility	No Growth	Pass
HSA Analysis	7.50 – 12.50 mg/mL	Pass
Gentamicin	0.0075 – 0.0125 mg/mL	Pass

Product passed and released by Quality

Signature:  Electronically signed by: Eila Cooper
Reason: QA Approval
Date: Sep 22, 2025 13:16:11 GMT+10

Name: Eila Cooper
Job Title: Quality Assurance Associate
Date: 22 Sep. 2025








Certificate of Analysis - J000007458

Final Audit Report

2025-09-22

Created:	2025-09-22 (Australian Eastern Standard Time)
By:	Eila Cooper (eila.cooper@geneabiomedx.com)
Status:	Signed
Transaction ID:	CBJCHBCAABAASiNAAzRaFZU4ZZ0TUT_xjjvrENyzzUO

"Certificate of Analysis - J000007458" History

-  Document created by Eila Cooper (eila.cooper@geneabiomedx.com)
2025-09-22 - 1:15:17 PM GMT+10- IP address: 163.47.70.120
-  Document emailed to Eila Cooper (eila.cooper@geneabiomedx.com) for signature
2025-09-22 - 1:15:29 PM GMT+10
-  Email viewed by Eila Cooper (eila.cooper@geneabiomedx.com)
2025-09-22 - 1:15:38 PM GMT+10- IP address: 68.218.165.129
-  Eila Cooper (eila.cooper@geneabiomedx.com) authenticated with Adobe Acrobat Sign.
Challenge: The user opened the agreement.
2025-09-22 - 1:15:47 PM GMT+10
-  Eila Cooper (eila.cooper@geneabiomedx.com) authenticated with Adobe Acrobat Sign.
Challenge: The user completed the signing ceremony.
2025-09-22 - 1:16:10 PM GMT+10
-  Document e-signed by Eila Cooper (eila.cooper@geneabiomedx.com)
Signing reason: QA Approval
Signature Date: 2025-09-22 - 1:16:11 PM GMT+10 - Time Source: server- IP address: 163.47.70.120
-  Agreement completed.
2025-09-22 - 1:16:11 PM GMT+10