

# Certificate of Analysis



Level 2, 321 Kent Street, Sydney NSW 2000, Australia

Phone: +61 2 8228 2719  
Email: [info@geneabiomedx.com](mailto:info@geneabiomedx.com)  
Website: [geneabiomedx.com](http://geneabiomedx.com)

## Product Details

**Product description:** Blastocyst Medium  
**Product code** BLM-50  
**Lot number:** J000007493  
**Expiry date:** 11 Aug. 2026

## Intended Use

Blastocyst Medium is used to provide necessary nutrients for embryo development *in vitro* from the cleavage stage to the blastocyst stage. The embryos may remain in this solution for up to 3 days prior to being transferred to the uterus.

**USA only:** Gems Blastocyst Medium is for *in vitro* culture of embryos from the cleavage stage to the blastocyst stage of development. This medium can be used for transfer of blastocyst stage embryos into the uterus.

**Caution:** Federal law (USA) restricts this device to sale by or on the order of a physician or practitioner trained in its use.

## Product Analysis

Assay	Specification	Result
pH at 37°C with 6 % CO <sub>2</sub>	7.25 – 7.45	Pass
Osmolality	285 – 295 mOsm/kg	Pass
MEA	MEA (1-cell): ≥80% developed to blastocyst stage or beyond by 96 hours	Pass
Endotoxin	<0.40 EU/mL	Pass
Sterility	No Growth	Pass
HSA Analysis	3.75 - 6.25 mg/mL	Pass
Gentamicin	0.0075 – 0.0125 mg/mL	Pass

## Product passed and released by Quality

**Signature:**  Electronically signed by: Eila Cooper  
Reason: QA Approval  
Date: Sep 25, 2025 08:56:45 GMT+10

**Name:** Eila Cooper  
**Job Title:** Quality Assurance Associate  
**Date:** 25 Sep. 2025








# Certificate of Analysis - J000007493

Final Audit Report

2025-09-25

Created:	2025-09-25 (Australian Eastern Standard Time)
By:	Eila Cooper (eila.cooper@geneabiomedx.com)
Status:	Signed
Transaction ID:	CBJCHBCAABAABAKkbPbFg5EBPsNpgsTn10XjbXwQC__Nzh

## "Certificate of Analysis - J000007493" History

-  Document created by Eila Cooper (eila.cooper@geneabiomedx.com)  
2025-09-25 - 8:55:53 AM GMT+10- IP address: 163.47.70.120
-  Document emailed to Eila Cooper (eila.cooper@geneabiomedx.com) for signature  
2025-09-25 - 8:56:07 AM GMT+10
-  Email viewed by Eila Cooper (eila.cooper@geneabiomedx.com)  
2025-09-25 - 8:56:15 AM GMT+10- IP address: 68.218.165.158
-  Eila Cooper (eila.cooper@geneabiomedx.com) authenticated with Adobe Acrobat Sign.  
Challenge: The user opened the agreement.  
2025-09-25 - 8:56:25 AM GMT+10
-  Eila Cooper (eila.cooper@geneabiomedx.com) authenticated with Adobe Acrobat Sign.  
Challenge: The user completed the signing ceremony.  
2025-09-25 - 8:56:44 AM GMT+10
-  Document e-signed by Eila Cooper (eila.cooper@geneabiomedx.com)  
Signing reason: QA Approval  
Signature Date: 2025-09-25 - 8:56:45 AM GMT+10 - Time Source: server- IP address: 163.47.70.120
-  Agreement completed.  
2025-09-25 - 8:56:45 AM GMT+10