Certificate of Analysis



Level 2, 321 Kent Street, Sydney NSW 2000, Australia

Phone: +61 2 8228 2719
Email: info@geneabiomedx.com
Website: geneabiomedx.com

Product Details

Product description: Sperm Buffer

Product code SPB-50

Lot number: J000007518

Expiry date: 24 Aug. 2026

Intended Use

Sperm Buffer is used to wash and resuspend sperm for the insemination step in intrauterine insemination (IUI), IVF or in diagnostic washing.

Product Analysis

| Assay | Specification | Result |
|---------------------|--|--------|
| pH at 25 °C in air | 7.3 – 7.5 | Pass |
| Osmolality | 285 – 295 mOsm/kg | Pass |
| MEA | MEA (1-cell): ≥80% developed to blastocyst stage or beyond by 96 hours | Pass |
| Sperm Toxicity Test | Sperm Survival Tested | Pass |
| Endotoxin | <0.40 EU/mL | Pass |
| Sterility | No Growth | Pass |
| HSA Analysis | 7.50 – 12.50 mg/mL | Pass |
| Gentamicin | 0.0075 – 0.0125 mg/mL | Pass |

Product passed and released by Quality

Signature: Electronically signed by: Eila Cooper
Reason: QA Approval
Date: Sep 25, 2025 14:55:50 GMT+10

Name: Eila Cooper

Job Title: Quality Assurance Associate

Date: 25 Sep. 2025

QFRM216-08 Page 1 of 1

Certificate of Analysis - J000007518

Final Audit Report 2025-09-25

Created: 2025-09-25 (Australian Eastern Standard Time)

By: Eila Cooper (eila.cooper@geneabiomedx.com)

Status: Signed

Transaction ID: CBJCHBCAABAA-EKsxWjXa4qhmuidlotbIHkljDTRptme

"Certificate of Analysis - J000007518" History

Document created by Eila Cooper (eila.cooper@geneabiomedx.com) 2025-09-25 - 2:54:57 PM GMT+10- IP address: 163.47.70.120

Document emailed to Eila Cooper (eila.cooper@geneabiomedx.com) for signature 2025-09-25 - 2:55:11 PM GMT+10

Email viewed by Eila Cooper (eila.cooper@geneabiomedx.com)

2025-09-25 - 2:55:19 PM GMT+10- IP address: 48.215.3.140

Eila Cooper (eila.cooper@geneabiomedx.com) authenticated with Adobe Acrobat Sign.

Challenge: The user opened the agreement.

2025-09-25 - 2:55:28 PM GMT+10

Document e-signed by Eila Cooper (eila.cooper@geneabiomedx.com)

Signing reason: QA Approval

Signature Date: 2025-09-25 - 2:55:50 PM GMT+10 - Time Source: server- IP address: 163.47.70.120

Agreement completed.

2025-09-25 - 2:55:50 PM GMT+10

Eila Cooper (eila.cooper@geneabiomedx.com) authenticated with Adobe Acrobat Sign.

Challenge: The user completed the signing ceremony.

2025-09-25 - 2:55:50 PM GMT+10

