

Certificate of Analysis



Level 2, 321 Kent Street, Sydney NSW 2000, Australia

Phone: +61 2 8228 2719
Email: info@geneabiomedx.com
Website: geneabiomedx.com

Product Details

Product description: Cleavage Medium
Product code CLM-50
Lot number: J000007534
Expiry date: 02 Sep. 2026

Intended Use

Cleavage Medium is used to provide necessary nutrients for embryo development *in vitro*. The embryos can remain in this solution for up to 2 days prior to being transferred to the uterus or grown for up to another 3 days in Blastocyst Medium (BLM-20, BLM-50).

USA only: Gems Cleavage Medium is for *in vitro* culture of embryos following fertilisation to the 4-8 cell stage. This medium can be used for transfer of cleavage stage embryos into the uterus.

Caution: Federal law (USA) restricts this device to sale by or on the order of a physician or practitioner trained in its use.

Product Analysis

Assay	Specification	Result
pH at 37 °C with 6% CO ₂	7.3 – 7.5	Pass
Osmolality	285 – 295 mOsm/kg	Pass
MEA	MEA (1-cell): ≥80% developed to blastocyst stage or beyond by 96 hours	Pass
Endotoxin	<0.40 EU/mL	Pass
Sterility	No Growth	Pass
HSA Analysis	3.75 – 6.25 mg/mL	Pass
Gentamicin	0.0075 – 0.0125 mg/mL	Pass

Product passed and released by Quality

Signature:  Electronically signed by: Eila Cooper
Reason: QA Approval
Date: Oct 1, 2025 15:01:28 GMT+10

Name: Eila Cooper
Job Title: Quality Assurance Associate
Date: 01 Oct. 2025

Certificate of Analysis - J000007534

Final Audit Report

2025-10-01

Created:	2025-10-01 (Australian Eastern Standard Time)
By:	Eila Cooper (eila.cooper@geneabiomedx.com)
Status:	Signed
Transaction ID:	CBJCHBCAABAAg5hs2LBOIwIBZlhHHgkyETK8hTI18gW

"Certificate of Analysis - J000007534" History

-  Document created by Eila Cooper (eila.cooper@geneabiomedx.com)
2025-10-01 - 3:00:01 PM GMT+10- IP address: 163.47.70.34
-  Document emailed to Eila Cooper (eila.cooper@geneabiomedx.com) for signature
2025-10-01 - 3:00:16 PM GMT+10
-  Email viewed by Eila Cooper (eila.cooper@geneabiomedx.com)
2025-10-01 - 3:00:26 PM GMT+10- IP address: 68.218.165.88
-  Eila Cooper (eila.cooper@geneabiomedx.com) authenticated with Adobe Acrobat Sign.
Challenge: The user opened the agreement.
2025-10-01 - 3:01:09 PM GMT+10
-  Document e-signed by Eila Cooper (eila.cooper@geneabiomedx.com)
Signing reason: QA Approval
Signature Date: 2025-10-01 - 3:01:28 PM GMT+10 - Time Source: server- IP address: 163.47.70.34
-  Agreement completed.
2025-10-01 - 3:01:28 PM GMT+10
-  Eila Cooper (eila.cooper@geneabiomedx.com) authenticated with Adobe Acrobat Sign.
Challenge: The user completed the signing ceremony.
2025-10-01 - 3:01:28 PM GMT+10