

Certificate of Analysis



Level 2, 321 Kent Street, Sydney NSW 2000, Australia

Phone: +61 2 8228 2719
Email: info@geneabiomedx.com
Website: geneabiomedx.com

Product Details

Product description: Vitrification Set
Product code VIT-01
Lot number: J000007545
Expiry date: 28 Sep. 2026

Intended Use

Gems Vitrification Set is used for the vitrification of human embryos for Assisted Reproductive Technology (ART) procedures.

Caution: Federal law (USA) restricts this device to sale by or on the order of a physician or practitioner trained in its use.

Set Contents

The Gems Vitrification set comprises of Vitsol 1, Vitsol 2 and Vitsol 3.

Product Analysis

Assay	Specification			Result
	Vitsol 1	Vitsol 2	Vitsol 3	
pH at 25 °C in air	7.3 – 7.5			Pass
MEA	MEA (1-cell): ≥80% developed to blastocyst stage or beyond by 96 hours			Pass
Endotoxin	<0.40 EU/mL		<0.05 EU/mL	Pass
Sterility	No Growth			Pass
HSA Analysis	13.70 – 22.90 mg/mL	12.20 – 20.30 mg/mL		Pass

Product passed and released by Quality

Signature:

Electronically signed by: Eila Cooper
Reason: QA Approval
Date: Oct 27, 2025 11:13:22 GMT+11

Name: Eila Cooper

Job Title: Quality Assurance Associate

Date: 27 Oct. 2025

Certificate of Analysis - J000007545

Final Audit Report

2025-10-27

Created:	2025-10-27 (Australian Eastern Daylight Time)
By:	Eila Cooper (eila.cooper@geneabiomedx.com)
Status:	Signed
Transaction ID:	CBJCHBCAABAAY7fvXKT7PKbDu6NX14YrmGOUQMYUWAZj

"Certificate of Analysis - J000007545" History

-  Document created by Eila Cooper (eila.cooper@geneabiomedx.com)
2025-10-27 - 11:12:19 AM GMT+11- IP address: 163.47.70.57
-  Document emailed to Eila Cooper (eila.cooper@geneabiomedx.com) for signature
2025-10-27 - 11:12:36 AM GMT+11
-  Email viewed by Eila Cooper (eila.cooper@geneabiomedx.com)
2025-10-27 - 11:12:46 AM GMT+11- IP address: 4.199.143.157
-  Eila Cooper (eila.cooper@geneabiomedx.com) authenticated with Adobe Acrobat Sign.
Challenge: The user opened the agreement.
2025-10-27 - 11:12:59 AM GMT+11
-  Eila Cooper (eila.cooper@geneabiomedx.com) authenticated with Adobe Acrobat Sign.
Challenge: The user completed the signing ceremony.
2025-10-27 - 11:13:21 AM GMT+11
-  Document e-signed by Eila Cooper (eila.cooper@geneabiomedx.com)
Signing reason: QA Approval
Signature Date: 2025-10-27 - 11:13:22 AM GMT+11 - Time Source: server- IP address: 163.47.70.57
-  Agreement completed.
2025-10-27 - 11:13:22 AM GMT+11