Certificate of Analysis



Level 2, 321 Kent Street, Sydney NSW 2000, Australia

Phone: +61 2 8228 2719
Email: info@geneabiomedx.com
Website: geneabiomedx.com

Product Details

Product description: Sperm Wash Gradient Set

Product code SWG-01
Lot number: J000007285
Expiry date: 24 Mar. 2026

Intended Use

Sperm Wash Gradient Set is used to separate sperm from seminal plasma as well as separating highly motile sperm in preparation for insemination.

Product Analysis

| Assay | Specification | | Result |
|---------------------|-----------------------|-------------------|--------|
| | 45% gradient | 90% gradient | Result |
| pH at 25 °C in air | 7.3 – 7.5 | | Pass |
| Osmolality | 285 - 295 mOsm/kg | 280 - 290 mOsm/kg | Pass |
| Sperm Toxicity Test | Sperm Survival Tested | | Pass |
| Endotoxin | <0.40 EU/mL | | Pass |
| Sterility | No Growth | | Pass |
| HSA Analysis | 7.50 – 12.50 mg/mL | | Pass |
| Gentamicin | 0.0075 – 0.0125 mg/mL | | Pass |

Product passed and released by Quality

Signature:

Electronically signed by: Nicole Connor Reason: QA Approval Date: Oct 29, 2025 10:00:20 GMT+11

Name: Nicole Connor

Job Title: Quality Assurance Associate

Date: 29 Oct. 2025

QFRM221-10 Page 1 of 1