

Certificate of Analysis



Level 2, 321 Kent Street, Sydney NSW 2000, Australia

Phone: +61 2 8228 2719

Email: info@geneabiomedx.com

Website: geneabiomedx.com

Product Details

Product description:	GERI Medium
Product code	ONE-50
Lot number:	J000007576
Expiry date:	07 Oct. 2026

Intended Use

Geri Medium is used to provide necessary nutrients for embryo development *in vitro*. The embryos can remain in this solution for up to 5 days prior to being transferred to the uterus.

USA only: Gems Geri Medium is for *in vitro* culture of embryos from fertilisation to the blastocyst stage of development. This medium can be used for transfer of embryos into the uterus.

Caution: Federal law (USA) restricts this device to sale by or on the order of a physician or practitioner trained in its use.

Product Analysis

Assay	Specification	Result
pH at 37 °C with 6 % CO ₂	7.3 – 7.5	Pass
Osmolality	285 – 295 mOsm/kg	Pass
MEA	MEA (1-cell): ≥80% developed to blastocyst stage or beyond by 96 hours	Pass
Endotoxin	<0.40 EU/mL	Pass
Sterility	No Growth	Pass
HSA Analysis	3.75 – 6.25 mg/mL	Pass
Gentamicin	0.0075 – 0.0125 mg/mL	Pass

Product passed and released by Quality

Signature:  Electronically signed by: Eila Cooper
Reason: QA Approval
Date: Oct 31, 2025 14:42:56 GMT+11

Name: Eila Cooper

Job Title: Quality Assurance Associate

Date: 31 Oct. 2025

Certificate of Analysis - J000007576

Final Audit Report

2025-10-31

Created:	2025-10-31 (Australian Eastern Daylight Time)
By:	Eila Cooper (eila.cooper@geneabiomedx.com)
Status:	Signed
Transaction ID:	CBJCHBCAABAA4WleB3IMgzIXv9wNPZAvUuDB1bw1OPwq

"Certificate of Analysis - J000007576" History

-  Document created by Eila Cooper (eila.cooper@geneabiomedx.com)
2025-10-31 - 2:41:55 PM GMT+11- IP address: 220.245.114.106
-  Document emailed to Eila Cooper (eila.cooper@geneabiomedx.com) for signature
2025-10-31 - 2:42:13 PM GMT+11
-  Email viewed by Eila Cooper (eila.cooper@geneabiomedx.com)
2025-10-31 - 2:42:20 PM GMT+11- IP address: 48.215.3.143
-  Eila Cooper (eila.cooper@geneabiomedx.com) authenticated with Adobe Acrobat Sign.
Challenge: The user opened the agreement.
2025-10-31 - 2:42:30 PM GMT+11
-  Document e-signed by Eila Cooper (eila.cooper@geneabiomedx.com)
Signing reason: QA Approval
Signature Date: 2025-10-31 - 2:42:56 PM GMT+11 - Time Source: server- IP address: 220.245.114.106
-  Agreement completed.
2025-10-31 - 2:42:56 PM GMT+11
-  Eila Cooper (eila.cooper@geneabiomedx.com) authenticated with Adobe Acrobat Sign.
Challenge: The user completed the signing ceremony.
2025-10-31 - 2:42:56 PM GMT+11