

Certificate of Analysis



Level 2, 321 Kent Street, Sydney NSW 2000, Australia

Phone: +61 2 8228 2719

Email: info@geneabiomedx.com

Website: geneabiomedx.com

Product Details

Product description:	VitBase
Product code	VBS-20
Lot number:	J000007662
Expiry date:	18 Nov. 2026

Intended Use

VitBase is a HEPES buffered medium that is used to maintain embryos for a short period of time in a non-gassed environment.

USA only: Gems VitBase is a HEPES buffered medium that is used to maintain embryos for a short period of time in a non-gassed environment during embryo vitrification and warming procedures. The embryos can be placed in this medium for a maximum of 10 minutes.

Caution: Federal law (USA) restricts this device to sale by or on the order of a physician or practitioner trained in its use.

Product Analysis

Assay	Specification	Result
pH at 25 °C in air	7.3 – 7.5	Pass
Osmolality	295 – 305 mOsm/kg	Pass
MEA	MEA (1-cell): ≥80% developed to blastocyst stage or beyond by 96 hours	Pass
Endotoxin	<0.40 EU/mL	Pass
Sterility	No Growth	Pass
HSA Analysis	15.00 – 25.00 mg/mL	Pass

Product passed and released by Quality

Signature:

A handwritten signature in black ink, appearing to read 'nconnor'.

Electronically signed by: Nicole Connor
Reason: QA Approval
Date: Dec 17, 2025 14:19:59 GMT+11

Name: Nicole Connor

Job Title: Quality Assurance Associate

Date: 17 Dec. 2025








Certificate of Analysis- J000007662

Final Audit Report

2025-12-17

Created:	2025-12-17 (Australian Eastern Daylight Time)
By:	Nicole Connor (nicole.connor@geneabiomedx.com)
Status:	Signed
Transaction ID:	CBJCHBCAABAAsck7iWhg-YH54_AaDxK_u7649-qdd78E

"Certificate of Analysis- J000007662" History

-  Document created by Nicole Connor (nicole.connor@geneabiomedx.com)
2025-12-17 - 2:18:04 PM GMT+11- IP address: 115.70.50.120
-  Document emailed to Nicole Connor (nicole.connor@geneabiomedx.com) for signature
2025-12-17 - 2:18:21 PM GMT+11
-  Email viewed by Nicole Connor (nicole.connor@geneabiomedx.com)
2025-12-17 - 2:18:28 PM GMT+11- IP address: 68.218.165.155
-  Nicole Connor (nicole.connor@geneabiomedx.com) authenticated with Adobe Acrobat Sign.
Challenge: The user opened the agreement.
2025-12-17 - 2:19:37 PM GMT+11
-  Nicole Connor (nicole.connor@geneabiomedx.com) authenticated with Adobe Acrobat Sign.
Challenge: The user completed the signing ceremony.
2025-12-17 - 2:19:58 PM GMT+11
-  Document e-signed by Nicole Connor (nicole.connor@geneabiomedx.com)
Signing reason: QA Approval
Signature Date: 2025-12-17 - 2:19:59 PM GMT+11 - Time Source: server- IP address: 115.70.50.120
-  Agreement completed.
2025-12-17 - 2:19:59 PM GMT+11