

# Certificate of Analysis



Level 2, 321 Kent Street, Sydney NSW 2000, Australia

Phone: +61 2 8228 2719

Email: [info@geneabiomedx.com](mailto:info@geneabiomedx.com)

Website: [geneabiomedx.com](http://geneabiomedx.com)

## Product Details

Product description:	Oocyte Retrieval Buffer
Product code	ORB-50
Lot number:	J000007700
Expiry date:	08 Dec. 2026

## Intended Use

Gems Oocyte Retrieval Buffer is used in the removal of oocytes from ovarian follicles.

**Caution:** Federal law (USA) restricts this device to sale by or on the order of a physician or practitioner trained in its use.

## Product Analysis

Assay	Specification	Result
pH at 25 °C in air	7.3 – 7.5	Pass
Osmolality	280 – 290 mOsm/kg	Pass
MEA	MEA (1-cell): ≥80% developed to blastocyst stage or beyond by 96 hours	Pass
Endotoxin	<0.40 EU/mL	Pass
Sterility	No Growth	Pass
Gentamicin	0.0075 – 0.0125 mg/mL	Pass

## Product passed and released by Quality

Signature:

A handwritten signature in black ink, appearing to read 'n Connor'.

Electronically signed by: Nicole Connor  
Reason: QA Approval  
Date: Jan 15, 2026 10:40:43 GMT+11

Name: Nicole Connor

Job Title: Quality Assurance Associate

Date: 15 Jan. 2026








# Certificate of Analysis- J000007700

Final Audit Report

2026-01-15

Created:	2026-01-15 (Australian Eastern Daylight Time)
By:	Nicole Connor (nicole.connor@geneabiomedx.com)
Status:	Signed
Transaction ID:	CBJCHBCAABAAb8j9Rb1ehlP-V9lhm-l_TYyuyZp_WyW

## "Certificate of Analysis- J000007700" History

-  Document created by Nicole Connor (nicole.connor@geneabiomedx.com)  
2026-01-15 - 10:38:12 AM GMT+11- IP address: 220.245.114.106
-  Document emailed to Nicole Connor (nicole.connor@geneabiomedx.com) for signature  
2026-01-15 - 10:38:30 AM GMT+11
-  Email viewed by Nicole Connor (nicole.connor@geneabiomedx.com)  
2026-01-15 - 10:38:38 AM GMT+11- IP address: 48.215.4.2
-  Nicole Connor (nicole.connor@geneabiomedx.com) authenticated with Adobe Acrobat Sign.  
Challenge: The user opened the agreement.  
2026-01-15 - 10:38:48 AM GMT+11
-  Document e-signed by Nicole Connor (nicole.connor@geneabiomedx.com)  
Signing reason: QA Approval  
Signature Date: 2026-01-15 - 10:40:43 AM GMT+11 - Time Source: server- IP address: 220.245.114.106
-  Agreement completed.  
2026-01-15 - 10:40:43 AM GMT+11
-  Nicole Connor (nicole.connor@geneabiomedx.com) authenticated with Adobe Acrobat Sign.  
Challenge: The user completed the signing ceremony.  
2026-01-15 - 10:40:43 AM GMT+11