

Certificate of Analysis



Level 2, 321 Kent Street, Sydney NSW 2000, Australia

Phone: +61 2 8228 2719
Email: info@geneabiomedx.com
Website: geneabiomedx.com

Product Details

Product description: Cleavage Medium
Product code: CLM-50
Lot number: J000007719
Expiry date: 12 Jan. 2027

Intended Use

Cleavage Medium is used to provide necessary nutrients for embryo development *in vitro*. The embryos can remain in this solution for up to 2 days prior to being transferred to the uterus or grown for up to another 3 days in Blastocyst Medium (BLM-20, BLM-50).

USA only: Gems Cleavage Medium is for *in vitro* culture of embryos following fertilisation to the 4-8 cell stage. This medium can be used for transfer of cleavage stage embryos into the uterus.

Caution: Federal law (USA) restricts this device to sale by or on the order of a physician or practitioner trained in its use.

Product Analysis

Assay	Specification	Result
pH at 37 °C with 6% CO ₂	7.3 – 7.5	Pass
Osmolality	285 – 295 mOsm/kg	Pass
MEA	MEA (1-cell): ≥80% developed to blastocyst stage or beyond by 96 hours	Pass
Endotoxin	<0.40 EU/mL	Pass
Sterility	No Growth	Pass
HSA Analysis	3.75 – 6.25 mg/mL	Pass
Gentamicin	0.0075 – 0.0125 mg/mL	Pass

Product passed and released by Quality

Signature:

Electronically signed by: Nicole Connor
Reason: QA Approval
Date: Feb 18, 2026 07:59:18 GMT +11

Name: Nicole Connor

Job Title: Quality Assurance Associate

Date: 18 Feb. 2026

Certificate of Analysis- J000007719

Final Audit Report

2026-02-18

Created:	2026-02-18 (Australian Eastern Daylight Time)
By:	Nicole Connor (nicole.connor@geneabiomedx.com)
Status:	Signed
Transaction ID:	CBJCHBCAABAAniuhnLN7aHEShKKk2ORWEwhkEYMWb8bO

"Certificate of Analysis- J000007719" History

-  Document created by Nicole Connor (nicole.connor@geneabiomedx.com)
2026-02-18 - 7:56:16 AM GMT+11- IP address: 149.167.36.126
-  Document emailed to Nicole Connor (nicole.connor@geneabiomedx.com) for signature
2026-02-18 - 7:58:37 AM GMT+11
-  Email viewed by Nicole Connor (nicole.connor@geneabiomedx.com)
2026-02-18 - 7:58:43 AM GMT+11- IP address: 4.199.143.141
-  Nicole Connor (nicole.connor@geneabiomedx.com) authenticated with Adobe Acrobat Sign.
Challenge: The user opened the agreement.
2026-02-18 - 7:58:56 AM GMT+11
-  Nicole Connor (nicole.connor@geneabiomedx.com) authenticated with Adobe Acrobat Sign.
Challenge: The user completed the signing ceremony.
2026-02-18 - 7:59:18 AM GMT+11
-  Document e-signed by Nicole Connor (nicole.connor@geneabiomedx.com)
Signing reason: QA Approval
Signature Date: 2026-02-18 - 7:59:18 AM GMT+11 - Time Source: server- IP address: 149.167.36.126
-  Agreement completed.
2026-02-18 - 7:59:18 AM GMT+11