

Certificate of Analysis



Level 2, 321 Kent Street, Sydney NSW 2000, Australia

Phone: +61 2 8228 2719
Email: info@geneabiomedx.com
Website: geneabiomedx.com

Product Details

Product description: Sperm Wash Gradient Set
Product code SWG-01
Lot number: J000007730
Expiry date: 18 Jan. 2027

Intended Use

Sperm Wash Gradient Set is used to separate sperm from seminal plasma as well as separating highly motile sperm in preparation for insemination.

Product Analysis

| Assay | Specification | | Result |
|---------------------|-----------------------|-------------------|--------|
| | 45% gradient | 90% gradient | |
| pH at 25 °C in air | 7.3 – 7.5 | | Pass |
| Osmolality | 285 - 295 mOsm/kg | 280 - 290 mOsm/kg | Pass |
| Sperm Toxicity Test | Sperm Survival Tested | | Pass |
| Endotoxin | <0.40 EU/mL | | Pass |
| Sterility | No Growth | | Pass |
| HSA Analysis | 7.50 – 12.50 mg/mL | | Pass |
| Gentamicin | 0.0075 – 0.0125 mg/mL | | Pass |

Product passed and released by Quality

Signature:

Electronically signed by: Nicole Connor
Reason: QA Approval
Date: Feb 24, 2026 14:28:38 GMT +11

Name: Nicole Connor

Job Title: Quality Assurance Associate

Date: 24 Feb. 2026

Certificate of Analysis- J000007730

Final Audit Report

2026-02-24

| | |
|-----------------|--|
| Created: | 2026-02-24 (Australian Eastern Daylight Time) |
| By: | Nicole Connor (nicole.connor@geneabiomedx.com) |
| Status: | Signed |
| Transaction ID: | CBJCHBCAABAA_H-PYap2SUX2YLzW8Unwuj0Eg8W0OZUY |

"Certificate of Analysis- J000007730" History

-  Document created by Nicole Connor (nicole.connor@geneabiomedx.com)
2026-02-24 - 2:27:34 PM GMT+11- IP address: 149.167.36.126
-  Document emailed to Nicole Connor (nicole.connor@geneabiomedx.com) for signature
2026-02-24 - 2:27:50 PM GMT+11
-  Email viewed by Nicole Connor (nicole.connor@geneabiomedx.com)
2026-02-24 - 2:27:57 PM GMT+11- IP address: 4.199.143.156
-  Nicole Connor (nicole.connor@geneabiomedx.com) authenticated with Adobe Acrobat Sign.
Challenge: The user opened the agreement.
2026-02-24 - 2:28:10 PM GMT+11
-  Nicole Connor (nicole.connor@geneabiomedx.com) authenticated with Adobe Acrobat Sign.
Challenge: The user completed the signing ceremony.
2026-02-24 - 2:28:37 PM GMT+11
-  Document e-signed by Nicole Connor (nicole.connor@geneabiomedx.com)
Signing reason: QA Approval
Signature Date: 2026-02-24 - 2:28:38 PM GMT+11 - Time Source: server- IP address: 149.167.36.126
-  Agreement completed.
2026-02-24 - 2:28:38 PM GMT+11