

# Certificate of Analysis



Level 2, 321 Kent Street, Sydney NSW 2000, Australia

Phone: +61 2 8228 2719  
Email: info@geneabiomedx.com  
Website: geneabiomedx.com

## Product Details

**Product description:** Fertilisation Medium  
**Product code:** FEM-50  
**Lot number:** J000007775  
**Expiry date:** 10 Feb. 2027

## Intended Use

Fertilisation Medium is used to provide a suitable environment for both oocytes and sperm, to promote optimal fertilisation rates.

**USA only:** Gems Fertilisation Medium is used to provide a suitable environment for both oocytes and sperm and support *in vitro* fertilisation. This medium can be used for transfer of zygotes into the uterus.

**Caution:** Federal law (USA) restricts this device to sale by or on the order of a physician or practitioner trained in its use.

## Product Analysis

Assay	Specification	Result
pH at 37°C with 6 % CO <sub>2</sub>	7.3 – 7.5	Pass
Osmolality	295 – 305 mOsm/kg	Pass
MEA	MEA (1-cell): ≥80% developed to blastocyst stage or beyond by 96 hours	Pass
Endotoxin	<0.40 EU/mL	Pass
Sterility	No Growth	Pass
HSA Analysis	3.75 - 6.25 mg/mL	Pass
Gentamicin	0.0075 – 0.0125 mg/mL	Pass

## Product passed and released by Quality

**Signature:**

Electronically signed by: Nicole Connor  
Reason: QA Approval  
Date: Mar 11, 2026 12:47:13 GMT  
+11

**Name:** Nicole Connor

**Job Title:** Quality Assurance Associate

**Date:** 11 Mar. 2026

# Certificate of Analysis- J000007775

Final Audit Report

2026-03-11

Created:	2026-03-11 (Australian Eastern Daylight Time)
By:	Nicole Connor (nicole.connor@geneabiomedx.com)
Status:	Signed
Transaction ID:	CBJCHBCAABAA51jiPfCEYeQK8Zf6a-cjWdefJbeOaK3R

## "Certificate of Analysis- J000007775" History

-  Document created by Nicole Connor (nicole.connor@geneabiomedx.com)  
2026-03-11 - 12:46:14 PM GMT+11- IP address: 220.245.114.106
-  Document emailed to Nicole Connor (nicole.connor@geneabiomedx.com) for signature  
2026-03-11 - 12:46:29 PM GMT+11
-  Email viewed by Nicole Connor (nicole.connor@geneabiomedx.com)  
2026-03-11 - 12:46:35 PM GMT+11- IP address: 48.215.4.2
-  Nicole Connor (nicole.connor@geneabiomedx.com) authenticated with Adobe Acrobat Sign.  
Challenge: The user opened the agreement.  
2026-03-11 - 12:46:45 PM GMT+11
-  Nicole Connor (nicole.connor@geneabiomedx.com) authenticated with Adobe Acrobat Sign.  
Challenge: The user completed the signing ceremony.  
2026-03-11 - 12:47:13 PM GMT+11
-  Document e-signed by Nicole Connor (nicole.connor@geneabiomedx.com)  
Signing reason: QA Approval  
Signature Date: 2026-03-11 - 12:47:13 PM GMT+11 - Time Source: server- IP address: 220.245.114.106
-  Agreement completed.  
2026-03-11 - 12:47:13 PM GMT+11