

Certificate of Analysis



Level 2, 321 Kent Street, Sydney NSW 2000, Australia

Phone: +61 2 8228 2719
Email: info@geneabiomedx.com
Website: geneabiomedx.com

Product Details

Product description: Vitrification Set
Product code VIT-01
Lot number: J000007737
Expiry date: 26 Jan. 2027

Intended Use

Gems Vitrification Set is used for the vitrification of human embryos for Assisted Reproductive Technology (ART) procedures.

Caution: Federal law (USA) restricts this device to sale by or on the order of a physician or practitioner trained in its use.

Set Contents

The Gems Vitrification set comprises of Vitsol 1, Vitsol 2 and Vitsol 3.

Product Analysis

Assay	Specification			Result
	Vitsol 1	Vitsol 2	Vitsol 3	
pH at 25 °C in air	7.3 – 7.5			Pass
MEA	MEA (1-cell): ≥80% developed to blastocyst stage or beyond by 96 hours			Pass
Endotoxin	<0.40 EU/mL		<0.05 EU/mL	Pass
Sterility	No Growth			Pass
HSA Analysis	13.70 – 22.90 mg/mL	12.20 – 20.30 mg/mL		Pass

Product passed and released by Quality

Signature:

Electronically signed by: Nicole Connor
Reason: QA Approval
Date: Mar 2, 2026 12:44:19 GMT +11

Name: Nicole Connor

Job Title: Quality Assurance Associate

Date: 02 Mar. 2026

Certificate of Analysis- J000007737

Final Audit Report

2026-03-02

Created:	2026-03-02 (Australian Eastern Daylight Time)
By:	Nicole Connor (nicole.connor@geneabiomedx.com)
Status:	Signed
Transaction ID:	CBJCHBCAABAAEeAEu9-Z_3JgY5H7ArkI1ReGJmW4Lv0M

"Certificate of Analysis- J000007737" History

-  Document created by Nicole Connor (nicole.connor@geneabiomedx.com)
2026-03-02 - 12:41:54 PM GMT+11- IP address: 115.70.50.178
-  Document emailed to Nicole Connor (nicole.connor@geneabiomedx.com) for signature
2026-03-02 - 12:42:15 PM GMT+11
-  Nicole Connor (nicole.connor@geneabiomedx.com) authenticated with Adobe Acrobat Sign.
Challenge: The user opened the agreement.
2026-03-02 - 12:42:32 PM GMT+11
-  Email viewed by Nicole Connor (nicole.connor@geneabiomedx.com)
2026-03-02 - 12:42:36 PM GMT+11- IP address: 4.199.143.130
-  Nicole Connor (nicole.connor@geneabiomedx.com) authenticated with Adobe Acrobat Sign.
Challenge: The user completed the signing ceremony.
2026-03-02 - 12:44:19 PM GMT+11
-  Document e-signed by Nicole Connor (nicole.connor@geneabiomedx.com)
Signing reason: QA Approval
Signature Date: 2026-03-02 - 12:44:19 PM GMT+11 - Time Source: server- IP address: 115.70.50.178
-  Agreement completed.
2026-03-02 - 12:44:19 PM GMT+11