

Certificate of Analysis



Level 2, 321 Kent Street, Sydney NSW 2000, Australia

Phone: +61 2 8228 2719
Email: info@geneabiomedx.com
Website: geneabiomedx.com

Product Details

Product description: Oocyte Retrieval Buffer
Product code: ORB-50
Lot number: J000007780
Expiry date: 15 Feb. 2027

Intended Use

Gems Oocyte Retrieval Buffer is used in the removal of oocytes from ovarian follicles.

Caution: Federal law (USA) restricts this device to sale by or on the order of a physician or practitioner trained in its use.

Product Analysis

Assay	Specification	Result
pH at 25 °C in air	7.3 – 7.5	Pass
Osmolality	280 – 290 mOsm/kg	Pass
MEA	MEA (1-cell): ≥80% developed to blastocyst stage or beyond by 96 hours	Pass
Endotoxin	<0.40 EU/mL	Pass
Sterility	No Growth	Pass
Gentamicin	0.0075 – 0.0125 mg/mL	Pass

Product passed and released by Quality

Signature:

Electronically signed by: Nicole Connor
Reason: QA Approval
Date: Mar 24, 2026 11:52:34 GMT
+11

Name: Nicole Connor

Job Title: Quality Assurance Associate

Date: 24 Mar. 2026







Certificate of Analysis- J000007780

Final Audit Report

2026-03-24

Created:	2026-03-24 (Australian Eastern Daylight Time)
By:	Nicole Connor (nicole.connor@geneabiomedx.com)
Status:	Signed
Transaction ID:	CBJCHBCAABAAW3VEN7fHC6sblcWHbb7Wi2GZ32QTDjvR

"Certificate of Analysis- J000007780" History

-  Document created by Nicole Connor (nicole.connor@geneabiomedx.com)
2026-03-24 - 11:51:20 AM GMT+11- IP address: 115.70.50.178
-  Document emailed to Nicole Connor (nicole.connor@geneabiomedx.com) for signature
2026-03-24 - 11:51:39 AM GMT+11
-  Nicole Connor (nicole.connor@geneabiomedx.com) authenticated with Adobe Acrobat Sign.
Challenge: The user opened the agreement.
2026-03-24 - 11:52:04 AM GMT+11
-  Nicole Connor (nicole.connor@geneabiomedx.com) authenticated with Adobe Acrobat Sign.
Challenge: The user completed the signing ceremony.
2026-03-24 - 11:52:33 AM GMT+11
-  Document e-signed by Nicole Connor (nicole.connor@geneabiomedx.com)
Signing reason: QA Approval
Signature Date: 2026-03-24 - 11:52:34 AM GMT+11 - Time Source: server- IP address: 115.70.50.178
-  Agreement completed.
2026-03-24 - 11:52:34 AM GMT+11