

# Certificate of Analysis



Level 2, 321 Kent Street, Sydney NSW 2000, Australia

Phone: +61 2 8228 2719  
Email: info@geneabiomedx.com  
Website: geneabiomedx.com

## Product Details

**Product description:** Vitrification Set  
**Product code** VIT-01  
**Lot number:** J000007764  
**Expiry date:** 15 Mar. 2027

## Intended Use

Gems Vitrification Set is used for the vitrification of human embryos for Assisted Reproductive Technology (ART) procedures.

**Caution:** Federal law (USA) restricts this device to sale by or on the order of a physician or practitioner trained in its use.

## Set Contents

The Gems Vitrification set comprises of Vitsol 1, Vitsol 2 and Vitsol 3.

## Product Analysis

Assay	Specification			Result
	Vitsol 1	Vitsol 2	Vitsol 3	
pH at 25 °C in air	7.3 – 7.5			Pass
MEA	MEA (1-cell): ≥80% developed to blastocyst stage or beyond by 96 hours			Pass
Endotoxin	<0.40 EU/mL		<0.05 EU/mL	Pass
Sterility	No Growth			Pass
HSA Analysis	13.70 – 22.90 mg/mL	12.20 – 20.30 mg/mL		Pass

## Product passed and released by Quality

**Signature:**

Electronically signed by: Nicole Connor  
Reason: QA Approval  
Date: Apr 30, 2026 07:54:08 GMT +10

**Name:** Nicole Connor

**Job Title:** Quality Assurance Associate

**Date:** 30 Apr. 2026








# Certificate of Analysis- J000007764

Final Audit Report

2026-04-30

Created:	2026-04-30 (Australian Eastern Standard Time)
By:	Nicole Connor (nicole.connor@geneabiomedx.com)
Status:	Signed
Transaction ID:	CBJCHBCAABAAaoxacOi9V3MfH7MVyxtSpYUqktHpQPB

## "Certificate of Analysis- J000007764" History

-  Document created by Nicole Connor (nicole.connor@geneabiomedx.com)  
2026-04-30 - 7:53:09 AM GMT+10- IP address: 220.245.114.106
-  Document emailed to Nicole Connor (nicole.connor@geneabiomedx.com) for signature  
2026-04-30 - 7:53:25 AM GMT+10
-  Email viewed by Nicole Connor (nicole.connor@geneabiomedx.com)  
2026-04-30 - 7:53:32 AM GMT+10- IP address: 68.218.165.131
-  Nicole Connor (nicole.connor@geneabiomedx.com) authenticated with Adobe Acrobat Sign.  
Challenge: The user opened the agreement.  
2026-04-30 - 7:53:42 AM GMT+10
-  Nicole Connor (nicole.connor@geneabiomedx.com) authenticated with Adobe Acrobat Sign.  
Challenge: The user completed the signing ceremony.  
2026-04-30 - 7:54:08 AM GMT+10
-  Document e-signed by Nicole Connor (nicole.connor@geneabiomedx.com)  
Signing reason: QA Approval  
Signature Date: 2026-04-30 - 7:54:08 AM GMT+10 - Time Source: server- IP address: 220.245.114.106
-  Agreement completed.  
2026-04-30 - 7:54:08 AM GMT+10