

Certificate of Analysis



Level 2, 321 Kent Street, Sydney NSW 2000, Australia

Phone: +61 2 8228 2719
Email: info@geneabiomedx.com
Website: geneabiomedx.com

Product Details

Product description: GERI Medium
Product code: ONE-50
Lot number: J000007820
Expiry date: 09 Mar. 2027

Intended Use

Geri Medium is used to provide necessary nutrients for embryo development *in vitro*. The embryos can remain in this solution for up to 5 days prior to being transferred to the uterus.

USA only: Gems Geri Medium is for *in vitro* culture of embryos from fertilisation to the blastocyst stage of development. This medium can be used for transfer of embryos into the uterus.

Caution: Federal law (USA) restricts this device to sale by or on the order of a physician or practitioner trained in its use.

Product Analysis

Assay	Specification	Result
pH at 37 °C with 6 % CO ₂	7.3 – 7.5	Pass
Osmolality	285 – 295 mOsm/kg	Pass
MEA	MEA (1-cell): ≥80% developed to blastocyst stage or beyond by 96 hours	Pass
Endotoxin	<0.40 EU/mL	Pass
Sterility	No Growth	Pass
HSA Analysis	3.75 – 6.25 mg/mL	Pass
Gentamicin	0.0075 – 0.0125 mg/mL	Pass

Product passed and released by Quality

Signature:  Electronically signed by: Nicole Connor
Reason: QA Approval
Date: Apr 21, 2026 09:45:44 GMT +10

Name: Nicole Connor
Job Title: Quality Assurance Associate
Date: 21 Apr. 2026








Certificate of Analysis- J000007820

Final Audit Report

2026-04-21

Created:	2026-04-21 (Australian Eastern Standard Time)
By:	Nicole Connor (nicole.connor@geneabiomedx.com)
Status:	Signed
Transaction ID:	CBJCHBCAABAA_87zNKFPvTnUkv9nRBdhLOVKX30UOEKF

"Certificate of Analysis- J000007820" History

-  Document created by Nicole Connor (nicole.connor@geneabiomedx.com)
2026-04-21 - 9:44:38 AM GMT+10- IP address: 220.245.114.106
-  Document emailed to Nicole Connor (nicole.connor@geneabiomedx.com) for signature
2026-04-21 - 9:44:53 AM GMT+10
-  Nicole Connor (nicole.connor@geneabiomedx.com) authenticated with Adobe Acrobat Sign.
Challenge: The user opened the agreement.
2026-04-21 - 9:45:11 AM GMT+10
-  Email viewed by Nicole Connor (nicole.connor@geneabiomedx.com)
2026-04-21 - 9:45:33 AM GMT+10- IP address: 48.215.3.143
-  Nicole Connor (nicole.connor@geneabiomedx.com) authenticated with Adobe Acrobat Sign.
Challenge: The user completed the signing ceremony.
2026-04-21 - 9:45:44 AM GMT+10
-  Document e-signed by Nicole Connor (nicole.connor@geneabiomedx.com)
Signing reason: QA Approval
Signature Date: 2026-04-21 - 9:45:44 AM GMT+10 - Time Source: server- IP address: 220.245.114.106
-  Agreement completed.
2026-04-21 - 9:45:44 AM GMT+10