

Certificate of Analysis



Level 2, 321 Kent Street, Sydney NSW 2000, Australia

Phone: +61 2 8228 2719
Email: info@geneabiomedx.com
Website: geneabiomedx.com

Product Details

Product description: Oocyte Retrieval Buffer
Product code: ORB-50
Lot number: J000007834
Expiry date: 23 Mar. 2027

Intended Use

Gems Oocyte Retrieval Buffer is used in the removal of oocytes from ovarian follicles.

Caution: Federal law (USA) restricts this device to sale by or on the order of a physician or practitioner trained in its use.

Product Analysis

Assay	Specification	Result
pH at 25 °C in air	7.3 – 7.5	Pass
Osmolality	280 – 290 mOsm/kg	Pass
MEA	MEA (1-cell): ≥80% developed to blastocyst stage or beyond by 96 hours	Pass
Endotoxin	<0.40 EU/mL	Pass
Sterility	No Growth	Pass
Gentamicin	0.0075 – 0.0125 mg/mL	Pass

Product passed and released by Quality

Signature:

Electronically signed by: Caitlin De Cruz
Reason: QA approval
Date: May 4, 2026 12:46:37 GMT +10

Name: Caitlin De Cruz

Job Title: Quality Assurance Associate

Date: 04 May. 2026








Certificate of Analysis- J000007834

Final Audit Report

2026-05-04

Created:	2026-05-04 (Australian Eastern Standard Time)
By:	Caitlin De Cruz (caitlin.deacruz@geneabiomedx.com)
Status:	Signed
Transaction ID:	CBJCHBCAABAAJf_gOt3CFLRdvApeUERU0VKHm0NqG_6u

"Certificate of Analysis- J000007834" History

-  Document created by Caitlin De Cruz (caitlin.deacruz@geneabiomedx.com)
2026-05-04 - 12:45:24 PM GMT+10- IP address: 111.220.55.245
-  Document emailed to Caitlin De Cruz (caitlin.deacruz@geneabiomedx.com) for signature
2026-05-04 - 12:45:45 PM GMT+10
-  Email viewed by Caitlin De Cruz (caitlin.deacruz@geneabiomedx.com)
2026-05-04 - 12:45:51 PM GMT+10- IP address: 68.218.165.147
-  Caitlin De Cruz (caitlin.deacruz@geneabiomedx.com) authenticated with Adobe Acrobat Sign.
Challenge: The user opened the agreement.
2026-05-04 - 12:46:08 PM GMT+10
-  Caitlin De Cruz (caitlin.deacruz@geneabiomedx.com) authenticated with Adobe Acrobat Sign.
Challenge: The user completed the signing ceremony.
2026-05-04 - 12:46:36 PM GMT+10
-  Document e-signed by Caitlin De Cruz (caitlin.deacruz@geneabiomedx.com)
Signing reason: QA approval
Signature Date: 2026-05-04 - 12:46:37 PM GMT+10 - Time Source: server- IP address: 111.220.55.245
-  Agreement completed.
2026-05-04 - 12:46:37 PM GMT+10