

Certificate of Analysis



Level 2, 321 Kent Street, Sydney NSW 2000, Australia

Phone: +61 2 8228 2719
Email: info@geneabiomedx.com
Website: geneabiomedx.com

Product Details

Product description: GERI Medium
Product code: ONE-50
Lot number: J000007827
Expiry date: 11 Mar. 2027

Intended Use

Geri Medium is used to provide necessary nutrients for embryo development *in vitro*. The embryos can remain in this solution for up to 5 days prior to being transferred to the uterus.


USA only: Gems Geri Medium is for *in vitro* culture of embryos from fertilisation to the blastocyst stage of development. This medium can be used for transfer of embryos into the uterus.

Caution: Federal law (USA) restricts this device to sale by or on the order of a physician or practitioner trained in its use.

Product Analysis

Assay	Specification	Result
pH at 37 °C with 6 % CO ₂	7.3 – 7.5	Pass
Osmolality	285 – 295 mOsm/kg	Pass
MEA	MEA (1-cell): ≥80% developed to blastocyst stage or beyond by 96 hours	Pass
Endotoxin	<0.40 EU/mL	Pass
Sterility	No Growth	Pass
HSA Analysis	3.75 – 6.25 mg/mL	Pass
Gentamicin	0.0075 – 0.0125 mg/mL	Pass

Product passed and released by Quality

Signature:  Electronically signed by: Caitlin De Cruz
Reason: QA approval
Date: Apr 29, 2026 12:23:18 GMT +10

Name: Caitlin De Cruz
Job Title: Quality Assurance Associate
Date: 29 Apr. 2026








Certificate of Analysis- J000007827

Final Audit Report

2026-04-29

Created:	2026-04-29 (Australian Eastern Standard Time)
By:	Caitlin De Cruz (caitlin.deacruz@geneabiomedx.com)
Status:	Signed
Transaction ID:	CBJCHBCAABAAU_767QbJPiselyGYWER5ZL1PY4UWzSBn

"Certificate of Analysis- J000007827" History

-  Document created by Caitlin De Cruz (caitlin.deacruz@geneabiomedx.com)
2026-04-29 - 12:21:25 PM GMT+10- IP address: 111.220.55.245
-  Document emailed to Caitlin De Cruz (caitlin.deacruz@geneabiomedx.com) for signature
2026-04-29 - 12:21:56 PM GMT+10
-  Email viewed by Caitlin De Cruz (caitlin.deacruz@geneabiomedx.com)
2026-04-29 - 12:22:07 PM GMT+10- IP address: 4.199.143.157
-  Caitlin De Cruz (caitlin.deacruz@geneabiomedx.com) authenticated with Adobe Acrobat Sign.
Challenge: The user opened the agreement.
2026-04-29 - 12:22:43 PM GMT+10
-  Caitlin De Cruz (caitlin.deacruz@geneabiomedx.com) authenticated with Adobe Acrobat Sign.
Challenge: The user completed the signing ceremony.
2026-04-29 - 12:23:17 PM GMT+10
-  Document e-signed by Caitlin De Cruz (caitlin.deacruz@geneabiomedx.com)
Signing reason: QA approval
Signature Date: 2026-04-29 - 12:23:18 PM GMT+10 - Time Source: server- IP address: 111.220.55.245
-  Agreement completed.
2026-04-29 - 12:23:18 PM GMT+10