

Certificate of Analysis



Level 2, 321 Kent Street, Sydney NSW 2000, Australia

Phone: +61 2 8228 2719
Email: info@geneabiomedx.com
Website: geneabiomedx.com

Product Details

Product description: GERI Medium
Product code: ONE-20
Lot number: J000007826
Expiry date: 11 Mar. 2027

Intended Use

Geri Medium is used to provide necessary nutrients for embryo development *in vitro*. The embryos can remain in this solution for up to 5 days prior to being transferred to the uterus.

USA only: Gems Geri Medium is for *in vitro* culture of embryos from fertilisation to the blastocyst stage of development. This medium can be used for transfer of embryos into the uterus.

Caution: Federal law (USA) restricts this device to sale by or on the order of a physician or practitioner trained in its use.

Product Analysis

Assay	Specification	Result
pH at 37 °C with 6 % CO ₂	7.3 – 7.5	Pass
Osmolality	285 – 295 mOsm/kg	Pass
MEA	MEA (1-cell): ≥80% developed to blastocyst stage or beyond by 96 hours	Pass
Endotoxin	<0.40 EU/mL	Pass
Sterility	No Growth	Pass
HSA Analysis	3.75 – 6.25 mg/mL	Pass
Gentamicin	0.0075 – 0.0125 mg/mL	Pass

Product passed and released by Quality

Signature:  Electronically signed by: Eila Cooper
Reason: QA Approval
Date: Apr 30, 2026 14:43:34 GMT
+10

Name: Eila Cooper
Job Title: Quality Assurance Associate
Date: 30 Apr. 2026







Certificate of Analysis- J000007826

Final Audit Report

2026-04-30

Created:	2026-04-30 (Australian Eastern Standard Time)
By:	Eila Cooper (eila.cooper@geneabiomedx.com)
Status:	Signed
Transaction ID:	CBJCHBCAABAACg21HaVxuksp656YJTBd3Zbf32v1Chh0

"Certificate of Analysis- J000007826" History

-  Document created by Eila Cooper (eila.cooper@geneabiomedx.com)
2026-04-30 - 2:42:24 PM GMT+10- IP address: 163.53.146.99
-  Document emailed to Eila Cooper (eila.cooper@geneabiomedx.com) for signature
2026-04-30 - 2:42:41 PM GMT+10
-  Email viewed by Eila Cooper (eila.cooper@geneabiomedx.com)
2026-04-30 - 2:42:47 PM GMT+10- IP address: 48.215.3.166
-  Eila Cooper (eila.cooper@geneabiomedx.com) authenticated with Adobe Acrobat Sign.
Challenge: The user opened the agreement.
2026-04-30 - 2:42:58 PM GMT+10
-  Eila Cooper (eila.cooper@geneabiomedx.com) authenticated with Adobe Acrobat Sign.
Challenge: The user completed the signing ceremony.
2026-04-30 - 2:43:33 PM GMT+10
-  Document e-signed by Eila Cooper (eila.cooper@geneabiomedx.com)
Signing reason: QA Approval
Signature Date: 2026-04-30 - 2:43:34 PM GMT+10 - Time Source: server- IP address: 163.53.146.99
-  Agreement completed.
2026-04-30 - 2:43:34 PM GMT+10