

# Certificate of Analysis



Level 2, 321 Kent Street, Sydney NSW 2000, Australia

Phone: +61 2 8228 2719  
Email: info@geneabiomedx.com  
Website: geneabiomedx.com

## Product Details

**Product description:** Sperm Medium  
**Product code:** SPM-50  
**Lot number:** J000007845  
**Expiry date:** 29 Mar. 2027

## Intended Use

Sperm Medium is used to wash and resuspend sperm for the insemination step in IUI, IVF or in diagnostic washing, optimised for storage in a 6% CO<sub>2</sub> incubator.

## Product Analysis

Assay	Specification	Result
pH at 37°C with 6 % CO <sub>2</sub>	7.3 – 7.5	Pass
Osmolality	285 – 295 mOsm/kg	Pass
MEA	MEA (1-cell): ≥80% developed to blastocyst stage or beyond by 96 hours	Pass
Sperm Toxicity Test	Sperm Survival Tested	Pass
Endotoxin	<0.40 EU/mL	Pass
Sterility	No Growth	Pass
HSA Analysis	7.50 – 12.50 mg/mL	Pass
Gentamicin	0.0075 – 0.0125 mg/mL	Pass

## Product passed and released by Quality

**Signature:**

Electronically signed by: Nicole Connor  
Reason: QA Approval  
Date: May 12, 2026 09:38:47 GMT +10

**Name:** Nicole Connor

**Job Title:** Quality Assurance Associate

**Date:** 12 May. 2026








# Certificate of Analysis- J000007845

Final Audit Report

2026-05-12

Created:	2026-05-12 (Australian Eastern Standard Time)
By:	Nicole Connor (nicole.connor@geneabiomedx.com)
Status:	Signed
Transaction ID:	CBJCHBCAABAAC2A5ijQpC6MX449bKplAo1lhkLZ2T7eF

## "Certificate of Analysis- J000007845" History

-  Document created by Nicole Connor (nicole.connor@geneabiomedx.com)  
2026-05-12 - 9:37:11 AM GMT+10- IP address: 220.245.114.106
-  Document emailed to Nicole Connor (nicole.connor@geneabiomedx.com) for signature  
2026-05-12 - 9:37:32 AM GMT+10
-  Nicole Connor (nicole.connor@geneabiomedx.com) authenticated with Adobe Acrobat Sign.  
Challenge: The user opened the agreement.  
2026-05-12 - 9:37:48 AM GMT+10
-  Email viewed by Nicole Connor (nicole.connor@geneabiomedx.com)  
2026-05-12 - 9:37:48 AM GMT+10- IP address: 4.199.143.157
-  Nicole Connor (nicole.connor@geneabiomedx.com) authenticated with Adobe Acrobat Sign.  
Challenge: The user completed the signing ceremony.  
2026-05-12 - 9:38:47 AM GMT+10
-  Document e-signed by Nicole Connor (nicole.connor@geneabiomedx.com)  
Signing reason: QA Approval  
Signature Date: 2026-05-12 - 9:38:47 AM GMT+10 - Time Source: server- IP address: 220.245.114.106 - Signature Appearance Selected: DRAW
-  Agreement completed.  
2026-05-12 - 9:38:47 AM GMT+10