

Certificate of Analysis



Level 2, 321 Kent Street, Sydney NSW 2000, Australia

Phone: +61 2 8228 2719
Email: info@geneabiomedx.com
Website: geneabiomedx.com

Product Details

Product description: Sperm Medium
Product code: SPM-20
Lot number: J000007844
Expiry date: 29 Mar. 2027

Intended Use

Sperm Medium is used to wash and resuspend sperm for the insemination step in IUI, IVF or in diagnostic washing, optimised for storage in a 6% CO₂ incubator.

Product Analysis

Assay	Specification	Result
pH at 37°C with 6 % CO ₂	7.3 – 7.5	Pass
Osmolality	285 – 295 mOsm/kg	Pass
MEA	MEA (1-cell): ≥80% developed to blastocyst stage or beyond by 96 hours	Pass
Sperm Toxicity Test	Sperm Survival Tested	Pass
Endotoxin	<0.40 EU/mL	Pass
Sterility	No Growth	Pass
HSA Analysis	7.50 – 12.50 mg/mL	Pass
Gentamicin	0.0075 – 0.0125 mg/mL	Pass

Product passed and released by Quality

Signature:

Electronically signed by: Nicole Connor
Reason: QA Approval
Date: May 8, 2026 10:31:32 GMT +10

Name: Nicole Connor

Job Title: Quality Assurance Associate

Date: 08 May. 2026








Certificate of Analysis- J000007844

Final Audit Report

2026-05-08

Created:	2026-05-08 (Australian Eastern Standard Time)
By:	Nicole Connor (nicole.connor@geneabiomedx.com)
Status:	Signed
Transaction ID:	CBJCHBCAABAAACe7Sgk-vyVOUFPxl3yDBSPPWwr3lz070

"Certificate of Analysis- J000007844" History

-  Document created by Nicole Connor (nicole.connor@geneabiomedx.com)
2026-05-08 - 10:29:45 AM GMT+10- IP address: 149.167.36.126
-  Document emailed to Nicole Connor (nicole.connor@geneabiomedx.com) for signature
2026-05-08 - 10:30:38 AM GMT+10
-  Email viewed by Nicole Connor (nicole.connor@geneabiomedx.com)
2026-05-08 - 10:30:45 AM GMT+10- IP address: 68.218.165.134
-  Nicole Connor (nicole.connor@geneabiomedx.com) authenticated with Adobe Acrobat Sign.
Challenge: The user opened the agreement.
2026-05-08 - 10:31:08 AM GMT+10
-  Nicole Connor (nicole.connor@geneabiomedx.com) authenticated with Adobe Acrobat Sign.
Challenge: The user completed the signing ceremony.
2026-05-08 - 10:31:32 AM GMT+10
-  Document e-signed by Nicole Connor (nicole.connor@geneabiomedx.com)
Signing reason: QA Approval
Signature Date: 2026-05-08 - 10:31:32 AM GMT+10 - Time Source: server- IP address: 149.167.36.126 - Signature Appearance Selected: DRAW
-  Agreement completed.
2026-05-08 - 10:31:32 AM GMT+10