

Certificate of Analysis



Level 2, 321 Kent Street, Sydney NSW 2000, Australia

Phone: +61 2 8228 2719
Email: info@geneabiomedx.com
Website: geneabiomedx.com

Product Details

Product description: Blastocyst Medium
Product code BLM-50
Lot number: J000007980
Expiry date: 03 Jun. 2027

Intended Use

Blastocyst Medium is used to provide necessary nutrients for embryo development *in vitro* from the cleavage stage to the blastocyst stage. The embryos may remain in this solution for up to 3 days prior to being transferred to the uterus.

USA only: Gems Blastocyst Medium is for *in vitro* culture of embryos from the cleavage stage to the blastocyst stage of development. This medium can be used for transfer of blastocyst stage embryos into the uterus.

Caution: Federal law (USA) restricts this device to sale by or on the order of a physician or practitioner trained in its use.

Product Analysis

Assay	Specification	Result
pH at 37°C with 6 % CO ₂	7.25 – 7.45	Pass
Osmolality	285 – 295 mOsm/kg	Pass
MEA	MEA (1-cell): ≥80% developed to blastocyst stage or beyond by 96 hours	Pass
Endotoxin	<0.40 EU/mL	Pass
Sterility	No Growth	Pass
HSA Analysis	3.75 - 6.25 mg/mL	Pass
Gentamicin	0.0075 – 0.0125 mg/mL	Pass

Product passed and released by Quality

Signature:

Electronically signed by: Nicole Connor
Reason: QA Approval
Date: Jul 7, 2026 13:58:06 GMT+10

Name: Nicole Connor

Job Title: Quality Assurance Associate

Date: 07 Jul. 2026

Certificate of Analysis- J000007980

Final Audit Report

2026-07-07

Created:	2026-07-07 (Australian Eastern Standard Time)
By:	Nicole Connor (nicole.connor@geneabiomedx.com)
Status:	Signed
Transaction ID:	CBJCHBCAABAA_bYjlyfSBtkWN4Rz-regXIOkn7FJltYF

"Certificate of Analysis- J000007980" History

-  Document created by Nicole Connor (nicole.connor@geneabiomedx.com)
2026-07-07 - 1:56:53 PM GMT+10- IP address: 220.245.114.106
-  Document emailed to Nicole Connor (nicole.connor@geneabiomedx.com) for signature
2026-07-07 - 1:57:11 PM GMT+10
-  Nicole Connor (nicole.connor@geneabiomedx.com) authenticated with Adobe Acrobat Sign.
Challenge: The user opened the agreement.
2026-07-07 - 1:57:29 PM GMT+10
-  Email viewed by Nicole Connor (nicole.connor@geneabiomedx.com)
2026-07-07 - 1:57:35 PM GMT+10- IP address: 4.199.143.157
-  Nicole Connor (nicole.connor@geneabiomedx.com) authenticated with Adobe Acrobat Sign.
Challenge: The user completed the signing ceremony.
2026-07-07 - 1:58:03 PM GMT+10
-  Document e-signed by Nicole Connor (nicole.connor@geneabiomedx.com)
Signing reason: QA Approval
Signature Date: 2026-07-07 - 1:58:06 PM GMT+10 - Time Source: server- IP address: 220.245.114.106 - Signature Appearance Selected: DRAW
-  Agreement completed.
2026-07-07 - 1:58:06 PM GMT+10