

Certificate of Analysis



Level 2, 321 Kent Street, Sydney NSW 2000, Australia

Phone: +61 2 8228 2719

Email: info@geneabiomedx.com

Website: geneabiomedx.com

Product Details

Product Description	GERI Dish
Product Code	GERI-DSH-20
Lot Number	GD0050-001
Manufacture Date	2026-03
Expiry date	2029-02

Product Specification

Requirement	Specification	Result
MEA	≥ 80% of 2PN's developed to Blastocyst stage by 96 hrs	Pass
Endotoxin	< 20EU/Device	Pass
Sterility	No Growth	Pass

This product complies with approved and released procedures within the Genea Biomedx EN/ISO 13485 compliant Quality Management System. This product and all its identified Part Numbers and Revisions conform to approved product specifications retained within the QMS. Where cited, lot numbers indicate traceable items within the QMS.

Black artefacts may be present in GERI Dish microwells. These artefacts are cosmetic in nature, and there is no associated impact on the intended use of the dish. The visual imperfection is purely aesthetic and does not pose a risk to product quality or safety. Please direct any queries to customerservices@geneabiomedx.com.

Product Passed and Released by Quality Assurance

Signature:

A handwritten signature in black ink that reads 'ECooper'.

Electronically signed by: Eila Cooper
Reason: RA Approval
Date: Jun 30, 2026 12:22:00 GMT
+10

Name: Eila Cooper

Job Title: Regulatory Affairs Associate

Date: 30 Jun. 2026

COA - GD0050-001

Final Audit Report

2026-06-30

Created:	2026-06-30 (Australian Eastern Standard Time)
By:	Eila Cooper (eila.cooper@geneabiomedx.com)
Status:	Signed
Transaction ID:	CBJCHBCAABAAU-_jvv0FexjwW8HyWoHDuq8tAMoqG5JK

"COA - GD0050-001" History

-  Document created by Eila Cooper (eila.cooper@geneabiomedx.com)
2026-06-30 - 12:19:36 PM GMT+10- IP address: 163.53.146.72
-  Document emailed to Eila Cooper (eila.cooper@geneabiomedx.com) for signature
2026-06-30 - 12:19:51 PM GMT+10
-  Email viewed by Eila Cooper (eila.cooper@geneabiomedx.com)
2026-06-30 - 12:20:06 PM GMT+10- IP address: 4.199.143.157
-  Eila Cooper (eila.cooper@geneabiomedx.com) authenticated with Adobe Acrobat Sign.
Challenge: The user opened the agreement.
2026-06-30 - 12:20:35 PM GMT+10
-  Eila Cooper (eila.cooper@geneabiomedx.com) authenticated with Adobe Acrobat Sign.
Challenge: The user completed the signing ceremony.
2026-06-30 - 12:21:57 PM GMT+10
-  Document e-signed by Eila Cooper (eila.cooper@geneabiomedx.com)
Signing reason: RA Approval
Signature Date: 2026-06-30 - 12:22:00 PM GMT+10 - Time Source: server- IP address: 163.53.146.72 - Signature Appearance Selected: IMAGE
-  Agreement completed.
2026-06-30 - 12:22:00 PM GMT+10