

Certificate of Analysis



Level 2, 321 Kent Street, Sydney NSW 2000, Australia

Phone: +61 2 8228 2719
Email: info@geneabiomedx.com
Website: geneabiomedx.com

Product Details

Product description: Sperm Buffer
Product code SPB-50
Lot number: J000007917
Expiry date: 10 May. 2027

Intended Use

Sperm Buffer is used to wash and resuspend sperm for the insemination step in intrauterine insemination (IUI), IVF or in diagnostic washing.

Product Analysis

Assay	Specification	Result
pH at 25 °C in air	7.3 – 7.5	Pass
Osmolality	285 – 295 mOsm/kg	Pass
MEA	MEA (1-cell): ≥80% developed to blastocyst stage or beyond by 96 hours	Pass
Sperm Toxicity Test	Sperm Survival Tested	Pass
Endotoxin	<0.40 EU/mL	Pass
Sterility	No Growth	Pass
HSA Analysis	7.50 – 12.50 mg/mL	Pass
Gentamicin	0.0075 – 0.0125 mg/mL	Pass

Product passed and released by Quality

Signature:

Electronically signed by: Caitlin De Cruz
Reason: QA approval
Date: Jun 10, 2026 13:54:58 GMT+10

Name: Caitlin De Cruz
Job Title: Quality Assurance Associate
Date: 10 Jun. 2026

Certificate of Analysis - J000007917

Final Audit Report

2026-06-10

Created:	2026-06-10 (Australian Eastern Standard Time)
By:	Caitlin De Cruz (caitlin.deacruz@geneabiomedx.com)
Status:	Signed
Transaction ID:	CBJCHBCAABAATBTYS80cpOiCxn1sl70o-J7GVUmRULyi

"Certificate of Analysis - J000007917" History

-  Document created by Caitlin De Cruz (caitlin.deacruz@geneabiomedx.com)
2026-06-10 - 1:53:18 PM GMT+10- IP address: 159.196.170.242
-  Document emailed to Caitlin De Cruz (caitlin.deacruz@geneabiomedx.com) for signature
2026-06-10 - 1:53:45 PM GMT+10
-  Email viewed by Caitlin De Cruz (caitlin.deacruz@geneabiomedx.com)
2026-06-10 - 1:53:53 PM GMT+10- IP address: 48.215.3.156
-  Caitlin De Cruz (caitlin.deacruz@geneabiomedx.com) authenticated with Adobe Acrobat Sign.
Challenge: The user opened the agreement.
2026-06-10 - 1:54:19 PM GMT+10
-  Caitlin De Cruz (caitlin.deacruz@geneabiomedx.com) authenticated with Adobe Acrobat Sign.
Challenge: The user completed the signing ceremony.
2026-06-10 - 1:54:57 PM GMT+10
-  Document e-signed by Caitlin De Cruz (caitlin.deacruz@geneabiomedx.com)
Signing reason: QA approval
Signature Date: 2026-06-10 - 1:54:58 PM GMT+10 - Time Source: server- IP address: 159.196.170.242 - Signature Appearance Selected: IMAGE
-  Agreement completed.
2026-06-10 - 1:54:58 PM GMT+10