Certificate of Analysis

Genea BIOMEDX
Level 2, 321 Kent Street, Sydney NSW 2000, Australia
Phone: +61 2 8228 2719
Email: info@geneabiomedx.com
Website: geneabiomedx.com

Product Details

Product description:	GERI Medium
Product code	ONE-50
Lot number:	J000007897
Expiry date:	29 Apr. 2027

Intended Use

Geri Medium is used to provide necessary nutrients for embryo development *in vitro*. The embryos can remain in this solution for up to 5 days prior to being transferred to the uterus.

USA only: Gems Geri Medium is for *in vitro* culture of embryos from fertilisation to the blastocyst stage of development. This medium can be used for transfer of embryos into the uterus.

Caution: Federal law (USA) restricts this device to sale by or on the order of a physician or practitioner trained in its use.

Product Analysis

Assay	Specification	Result
pH at 37 °C with 6 % CO_2	7.3 – 7.5	**Pass**
Osmolality	285 – 295 mOsm/kg	**Pass**
MEA	MEA (1-cell): ≥80% developed to blastocyst stage or beyond by 96 hours	**Pass**
Endotoxin	<0.40 EU/mL	**Pass**
Sterility	No Growth	**Pass**
HSA Analysis	3.75 – 6.25 mg/mL	**Pass**
Gentamicin	0.0075 – 0.0125 mg/mL	**Pass**

Product passed and released by Quality

Signature: Electronically signed by: Caitlin De Cruz
Reason: QA approval
Date: Jun 9, 2026 14:34:49 GMT +10

Name: Caitlin De Cruz

Job Title: Quality Assurance Associate

Date: 09 Jun. 2026

Certificate of Analysis- J000007897

Final Audit Report 2026-06-09

Created:	2026-06-09 (Australian Eastern Standard Time)
By:	Caitlin De Cruz (caitlin.decruz@geneabiomedx.com)
Status:	Signed
Transaction ID:	CBJCHBCAABAAUI64u8BEhkA_5Ru-Ls8UwFOBnRuB3gvM

"Certificate of Analysis- J000007897" History

Document created by Caitlin De Cruz (caitlin.decruz@geneabiomedx.com)
2026-06-09 - 2:31:58 PM GMT+10- IP address: 159.196.170.242

Document emailed to Caitlin De Cruz (caitlin.decruz@geneabiomedx.com) for signature
2026-06-09 - 2:33:46 PM GMT+10

Email viewed by Caitlin De Cruz (caitlin.decruz@geneabiomedx.com)
2026-06-09 - 2:33:52 PM GMT+10- IP address: 48.215.3.140

Caitlin De Cruz (caitlin.decruz@geneabiomedx.com) authenticated with Adobe Acrobat Sign.
Challenge: The user opened the agreement.
2026-06-09 - 2:34:13 PM GMT+10

Caitlin De Cruz (caitlin.decruz@geneabiomedx.com) authenticated with Adobe Acrobat Sign.
Challenge: The user completed the signing ceremony.
2026-06-09 - 2:34:48 PM GMT+10

Document e-signed by Caitlin De Cruz (caitlin.decruz@geneabiomedx.com)
Signing reason: QA approval
Signature Date: 2026-06-09 - 2:34:49 PM GMT+10 - Time Source: server- IP address: 159.196.170.242 - Signature Appearance Selected: IMAGE

Agreement completed.
2026-06-09 - 2:34:49 PM GMT+10